

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB135586

Slope : 98.6

BalanceID: WC SC-7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3178
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3191
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3161
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	04/29/2025	08:20
2	CAL2	1	Water	NA	NA	20.2	7.00	04/29/2025	08:21
3	CAL3	1	Water	NA	NA	20.2	10.02	04/29/2025	08:22
4	ICV	1	Water	NA	NA	20.3	7.01	04/29/2025	08:25
5	CCV1	1	Water	NA	NA	20.2	2.01	04/29/2025	08:30
6	Q1901-02	1	Solid	20.02	20	21.3	8.45	04/29/2025	08:35
7	Q1901-03	1	Solid	20.03	20	20.9	8.99	04/29/2025	08:40
8	Q1901-04	1	Solid	20.02	20	21.0	7.07	04/29/2025	08:44
9	Q1901-05	1	Solid	20.03	20	20.7	8.16	04/29/2025	08:47
10	Q1901-05DUP	1	Solid	20.4	20	20.8	8.17	04/29/2025	08:48
11	CCV2	1	Water	NA	NA	20.2	12.02	04/29/2025	09:35

WORKLIST(Hardcopy Internal Chain)

JB 135586

WorkList Name : PH Q1907 WorkList ID : 189191 Department : Wet-Chemistry Date : 04-29-2025 08:00:44

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1901-02	B-167-SB01	Solid	pH	Cool 4 deg C	PORT06	L51	04/26/2025	9045D
Q1901-03	B-170-SB01	Solid	pH	Cool 4 deg C	PORT06	L51	04/26/2025	9045D
Q1901-04	B-167-SB02	Solid	pH	Cool 4 deg C	PORT06	L51	04/26/2025	9045D
Q1901-05	B-170-SB02	Solid	pH	Cool 4 deg C	PORT06	L51	04/26/2025	9045D

Date/Time 04/29/25 08:15
 Raw Sample Received by: JB
 Raw Sample Relinquished by: JP

Date/Time 04/29/25
 Raw Sample Received by: JP
 Raw Sample Relinquished by: JP