

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Iwona

Parameter: Flash Point

Supervisor Review By: jignesh

Run Number: LB135590

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3193

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.1	04/29/2025	09:30
2	Q1865-06		1	6	55.00	131.0	130.7	04/29/2025	10:00
3	Q1875-05		1	12	100.00	>212.0	>212.0	04/29/2025	10:30
4	Q1875-06		1	12	100.00	>212.0	>212.0	04/29/2025	11:00
5	Q1875-07		1	13	100.00	>212.0	>212.0	04/29/2025	11:30
6	Q1875-08		1	12	100.00	>212.0	>212.0	04/29/2025	12:00
7	Q1875-08DUP		1	14	100.00	>212.0	>212.0	04/29/2025	12:30

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB135590

WorkList Name : f04-29

WorkList ID : 189209

Department : Wet-Chemistry

Date : 04-29-2025 09:14:41

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1865-06	SOM-25-00017	Water	Flash Point	Cool 4 deg C	PSEG03	L41	04/23/2025	1010B
Q1875-05	AUD-25-0051	Water	Flash Point	Cool 4 deg C	PSEG03	L41	04/24/2025	1010B
Q1875-06	AUD-25-0052	Water	Flash Point	Cool 4 deg C	PSEG03	L41	04/24/2025	1010B
Q1875-07	AUD-25-0055	Water	Flash Point	Cool 4 deg C	PSEG03	L41	04/24/2025	1010B
Q1875-08	AUD-25-0056	Water	Flash Point	Cool 4 deg C	PSEG03	L41	04/24/2025	1010B

Date/Time

Raw Sample Received by:

Raw Sample Relinquished by:

04/29/25 09:20

12 (SC)

re/aa

Date/Time

Raw Sample Received by:

Raw Sample Relinquished by:

04/29/25

13:10

re/aa

12 (SC)