

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Iwona

Parameter: Flash Point

Supervisor Review By: jignesh

Run Number: LB135602

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3193

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	27.00	80.6	80.6	04/30/2025	09:30
2	Q1890-01		1	13	100.00	>212.0	>212.0	04/30/2025	10:00
3	Q1890-01DUP		1	14	100.00	>212.0	>212.0	04/30/2025	10:30
4	Q1896-03		1	14	100.00	>212.0	>212.0	04/30/2025	11:00

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

16135602

WorkList Name : FLASH POINT-4-29 WorkList ID : 189216 Department : Wet-Chemistry Date : 04-29-2025 16:11:16

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1890-01	TW-WTS-07	Water	Flash Point	Cool 4 deg C	ENTA05	L41	04/24/2025	1010B
Q1896-03	295-BERGEN-FRAC	Water	Flash Point	Cool 4 deg C	PSEG03	L31	04/25/2025	1010B

Date/Time 04/30/25 09:20
Raw Sample Received by: 12(50)
Raw Sample Relinquished by: JACOB

Date/Time 04/30/25 11:40
Raw Sample Received by: JACOB
Raw Sample Relinquished by: 12(50)