

L135611


7471B

INSTRUMENT ID: EV1

Linear

μ Abs.:

55944



A= 0.0000e+000

B= 1.7969e-004

C= 9.6727e-003

Rho= 0.9999368

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.020	0.020	57	0.000	57					1
0.2	0.200	0.217	0.017	1153	0.0 %	1153					8
2.5	2.500	2.508	0.008	13903	0.0 %	13903					0
5.0	5.000	4.942	-0.058	27447	0.0 %	27447					1
7.5	7.500	7.452	-0.048	41418	0.0 %	41418					1
10.0	10.000	10.062	0.062	55944	0.0 %	55944					1

LB135611 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 30 Apr 2025 13:02:46

Sample ID	Extended ID	μ Abs	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	57	-	0.0000	7471B	PPB	30 Apr 2025 13:17:20	S	Std
0.2 - 1	51.2	1153	-	0.2000	7471B	PPB	30 Apr 2025 13:19:37	S	Std
2.5 - 1	52.5	13903	-	2.5000	7471B	PPB	30 Apr 2025 13:21:57	S	Std
5.0 - 1	55	27447	-	5.0000	7471B	PPB	30 Apr 2025 13:24:14	S	Std
7.5 - 1	57.5	41418	-	7.5000	7471B	PPB	30 Apr 2025 13:26:33	S	Std
10.0 - 1	510	55944	-	10.0000	7471B	PPB	30 Apr 2025 13:28:53	S	Std
ICV109 - 1	ICV109	23346	4.2046	-	7471B	PPB	30 Apr 2025 13:35:54	U	SMPL
ICB109 - 1	ICB109	-342	-0.0518	-	7471B	PPB	30 Apr 2025 13:38:10	U	SMPL
CCV52 - 1	CCV52	27889	5.0209	-	7471B	PPB	30 Apr 2025 13:40:28	U	SMPL
CCB52 - 1	CCB52	-390	-0.0604	-	7471B	PPB	30 Apr 2025 13:42:42	U	SMPL
CRA - 1	CRA	1125	0.2118	-	7471B	PPB	30 Apr 2025 13:45:02	U	SMPL
HighStd - 1	HighStd	56588	10.1777	-	7471B	PPB	30 Apr 2025 13:47:17	U	SMPL
ChkStd - 1	ChkStd	37687	6.7815	-	7471B	PPB	30 Apr 2025 13:49:32	U	SMPL
PB167786BL - 1	PBS	-610	-0.0999	-	7471B	PPB	30 Apr 2025 13:51:56	U	SMPL
PB167786BS - 1	LCSS	19620	3.6351	-	7471B	PPB	30 Apr 2025 13:56:59	U	SMPL
Q1901-02 - 1	B-167-SB01	2696	0.4941	-	7471B	PPB	30 Apr 2025 13:59:16	U	SMPL
Q1901-02DUP - 1	B-167-SB01DUP	3070	0.5613	-	7471B	PPB	30 Apr 2025 14:01:36	U	SMPL
Q1901-02MS - 1	B-167-SB01MS	22870	4.1191	-	7471B	PPB	30 Apr 2025 14:03:55	U	SMPL
Q1901-02MSD - 1	B-167-SB01MSD	23071	4.1552	-	7471B	PPB	30 Apr 2025 14:06:13	U	SMPL
Q1901-03 - 1	B-170-SB01	16555	2.9844	-	7471B	PPB	30 Apr 2025 14:08:33	U	SMPL
CCV53 - 1	CCV53	27711	4.9889	-	7471B	PPB	30 Apr 2025 14:10:51	U	SMPL
CCB53 - 1	CCB53	-447	-0.0706	-	7471B	PPB	30 Apr 2025 14:13:10	U	SMPL
Q1901-04 - 1	B-167-SB02	3422	0.6246	-	7471B	PPB	30 Apr 2025 14:15:30	U	SMPL
Q1901-05 - 1	B-170-SB02	1174	0.2206	-	7471B	PPB	30 Apr 2025 14:17:45	U	SMPL
Q1902-02 - 1	343	2711	0.4968	-	7471B	PPB	30 Apr 2025 14:20:02	U	SMPL
Q1903-01 - 1	COMP-4	1173	0.2204	-	7471B	PPB	30 Apr 2025 14:22:18	U	SMPL
Q1903-02 - 1	COMP-5	1204	0.2260	-	7471B	PPB	30 Apr 2025 14:24:34	U	SMPL
Q1903-03 - 1	COMP-6	2577	0.4727	-	7471B	PPB	30 Apr 2025 14:26:51	U	SMPL
Q1904-01 - 1	VNJ-210	1320	0.2469	-	7471B	PPB	30 Apr 2025 14:29:07	U	SMPL
Q1905-01 - 1	MH-G	664	0.1290	-	7471B	PPB	30 Apr 2025 14:31:24	U	SMPL
Q1905-05 - 1	MH-H	1112	0.2095	-	7471B	PPB	30 Apr 2025 14:33:41	U	SMPL
Q1906-01 - 1	WC-4	141	0.0350	-	7471B	PPB	30 Apr 2025 14:35:57	U	SMPL
CCV54 - 1	CCV54	27862	5.0161	-	7471B	PPB	30 Apr 2025 14:38:12	U	SMPL
CCB54 - 1	CCB54	-361	-0.0552	-	7471B	PPB	30 Apr 2025 14:40:28	U	SMPL
Q1906-05 - 1	WC-5	7598	1.3749	-	7471B	PPB	30 Apr 2025 14:42:46	U	SMPL
Q1906-09 - 1	WC-6	7398	1.3390	-	7471B	PPB	30 Apr 2025 14:45:02	U	SMPL
Q1906-13 - 1	WC-7	1197	0.2248	-	7471B	PPB	30 Apr 2025 14:47:20	U	SMPL
Q1907-01 - 1	CO-8R-WC	23951	4.3133	-	7471B	PPB	30 Apr 2025 14:49:38	U	SMPL
Q1901-02LX5 - 1		161	0.0386	-	7471B	PPB	30 Apr 2025 14:51:53	U	SMPL
Q1901-02A - 1		24431	4.3996	-	7471B	PPB	30 Apr 2025 14:54:14	U	SMPL
CCV55 - 1	CCV55	28098	5.0585	-	7471B	PPB	30 Apr 2025 14:56:31	U	SMPL
CCB55 - 1	CCB55	-342	-0.0518	-	7471B	PPB	30 Apr 2025 14:58:51	U	SMPL