

LB135616


7471B

INSTRUMENT ID: CVL

Linear

μ Abs.:

49986



A= 0.0000e+000

B= 1.9960e-004

C= 2.1535e-003

Rho= 0.9998093

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	Q/D
0.0	0.000	-0.004	-0.004	-29	0.000	-29					1
0.2	0.200	0.247	0.047	1227	0.0 %	1227					24
2.5	2.500	2.376	-0.124	11892	0.0 %	11892					15
5.0	5.000	5.114	0.114	25609	0.0 %	25609					2
7.5	7.500	7.488	-0.012	37502	0.0 %	37502					00
10.0	10.000	9.979	-0.021	49986	0.0 %	49986					0

LB135616 INSTRUMENTY ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 01 May 2025 08:57:53

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-29	-	0.0000	7471B	PPB	01 May 2025 09:27:20	S	Std
0.2 - 1	50.2	1227	-	0.2000	7471B	PPB	01 May 2025 09:29:36	S	Std
2.5 - 1	52.5	11892	-	2.5000	7471B	PPB	01 May 2025 09:31:54	S	Std
5.0 - 1	55	25609	-	5.0000	7471B	PPB	01 May 2025 09:34:11	S	Std
7.5 - 1	57.5	37502	-	7.5000	7471B	PPB	01 May 2025 09:36:30	S	Std
10.0 - 1	57.5	49986	-	10.0000	7471B	PPB	01 May 2025 09:38:48	S	Std
ICV111 - 1	ICV111	20042	4.0026	-7471B	PPB	01 May 2025 09:42:45	U	SMPL	
ICB111 - 1	ICB111	-374	-0.0725	-7471B	PPB	01 May 2025 09:45:00	U	SMPL	
CCV59 - 1	CCV59	26642	5.3199	-7471B	PPB	01 May 2025 09:47:17	U	SMPL	
CCB59 - 1	CCB59	-359	-0.0695	-7471B	PPB	01 May 2025 09:49:33	U	SMPL	
CRA - 1	CRA	1238	0.2493	-7471B	PPB	01 May 2025 09:51:53	U	SMPL	
HighStd - 1	HighStd	50745	10.1309	-7471B	PPB	01 May 2025 09:54:08	U	SMPL	
ChkStd - 1	ChkStd	36007	7.1892	-7471B	PPB	01 May 2025 09:56:24	U	SMPL	
PB167814BL - 1	PBS	-65	-0.0108	-7471B	PPB	01 May 2025 10:01:38	U	SMPL	
PB167814BS - 1	LCSS	20412	4.0764	-7471B	PPB	01 May 2025 10:05:56	U	SMPL	
Q1910-01 - 1	MOO-25-0123-27	6765	1.3525	-7471B	PPB	01 May 2025 10:12:59	U	SMPL	
Q1910-01DUP - 1	MOO-25-0123-27DUP	5668	1.1335	-7471B	PPB	01 May 2025 10:15:18	U	SMPL	
Q1910-01MS - 1	MOO-25-0123-27MS	27839	5.5589	-7471B	PPB	01 May 2025 10:17:37	U	SMPL	
Q1910-01MSD - 1	MOO-25-0123-27MSD	30843	6.1585	-7471B	PPB	01 May 2025 10:19:55	U	SMPL	
Q1911-01 - 1	EO-03-04292025	13742	2.7451	-7471B	PPB	01 May 2025 10:22:14	U	SMPL	
CCV60 - 1	CCV60	26301	5.2519	-7471B	PPB	01 May 2025 10:24:32	U	SMPL	
CCB60 - 1	CCB60	-500	-0.0976	-7471B	PPB	01 May 2025 10:26:50	U	SMPL	
Q1912-01 - 1	MH-E	1865	0.3744	-7471B	PPB	01 May 2025 10:29:09	U	SMPL	
Q1912-05 - 1	MH-F	1147	0.2311	-7471B	PPB	01 May 2025 10:31:25	U	SMPL	
Q1916-01 - 1	WC-12	1549	0.3113	-7471B	PPB	01 May 2025 10:33:42	U	SMPL	
Q1917-01 - 1	MH-JJ	1111	0.2239	-7471B	PPB	01 May 2025 10:35:58	U	SMPL	
Q1920-01 - 1	291	11572	2.3119	-7471B	PPB	01 May 2025 10:38:22	U	SMPL	
Q1922-01 - 1	MH-R	961	0.1940	-7471B	PPB	01 May 2025 10:40:38	U	SMPL	
Q1922-05 - 1	MH-S	871	0.1760	-7471B	PPB	01 May 2025 10:42:57	U	SMPL	
Q1923-01 - 1	TR-05-04302025	1790	0.3594	-7471B	PPB	01 May 2025 10:45:14	U	SMPL	
Q1910-01LX5 - 1		968	0.1954	-7471B	PPB	01 May 2025 10:47:32	U	SMPL	
Q1910-01A - 1		30516	6.0932	-7471B	PPB	01 May 2025 10:49:47	U	SMPL	
CCV61 - 1	CCV61	27043	5.4000	-7471B	PPB	01 May 2025 10:52:03	U	SMPL	
CCB61 - 1	CCB61	-422	-0.0821	-7471B	PPB	01 May 2025 10:54:22	U	SMPL	