

Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB135619

Analyst By : jignesh

Supervisor Review By : Iwona

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3178
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3191
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3161
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	05/01/2025	08:15
2	CAL2	1	Water	NA	NA	20.3	7.00	05/01/2025	08:16
3	CAL3	1	Water	NA	NA	20.3	10.02	05/01/2025	08:20
4	ICV	1	Water	NA	NA	20.3	7.01	05/01/2025	08:27
5	CCV1	1	Water	NA	NA	20.2	2.01	05/01/2025	08:33
6	Q1919-03	1	Water	NA	NA	23.4	8.35	05/01/2025	08:40
7	Q1919-03DUP	1	Water	NA	NA	23.5	8.36	05/01/2025	08:41
8	Q1919-04	1	Water	NA	NA	23.7	7.80	05/01/2025	08:48
9	Q1919-06	1	Water	NA	NA	23.8	7.66	05/01/2025	08:55
10	Q1919-07	1	Water	NA	NA	24.1	7.69	05/01/2025	09:00
11	Q1925-04	1	Water	NA	NA	20.4	5.51	05/01/2025	09:11
12	Q1925-05	1	Water	NA	NA	20.6	4.80	05/01/2025	09:20
13	Q1926-01	1	Water	NA	NA	21.9	7.37	05/01/2025	09:30
14	CCV2	1	Water	NA	NA	20.2	12.02	05/01/2025	09:35

WORKLIST(Hardcopy Internal Chain)

1335619

WorkList Name : PH W Q1926 WorkList ID : 189244 Department : Wet-Chemistry Date : 05-01-2025 07:54:43

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1919-03	0423-A	Water	pH	Cool 4 deg C	PSEG03	L41	04/30/2025	9040C
Q1919-04	0423-B	Water	pH	Cool 4 deg C	PSEG03	L41	04/30/2025	9040C
Q1919-06	0428-A	Water	pH	Cool 4 deg C	PSEG03	L41	04/30/2025	9040C
Q1919-07	0428-B	Water	pH	Cool 4 deg C	PSEG03	L41	04/30/2025	9040C
Q1925-04	AUD-25-0072	Water	pH	Cool 4 deg C	PSEG03	L31	04/30/2025	9040C
Q1925-05	AUD-25-0073	Water	pH	Cool 4 deg C	PSEG03	L31	04/30/2025	9040C
Q1926-01	TRE-1104	Water	pH	Cool 4 deg C	PSEG03	L31	04/30/2025	9040C

Date/Time 05/01/25 08:00
 Raw Sample Received by: JH CWC
 Raw Sample Relinquished by: JH CWC

Date/Time 05/01/25 12:30
 Raw Sample Received by: JH CWC
 Raw Sample Relinquished by: JH CWC