

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB135620

Slope : 98.6

BalanceID: WC SC-7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3178
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3191
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3161
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	05/01/2025	09:30
2	CAL2	1	Water	NA	NA	20.2	7.00	05/01/2025	09:31
3	CAL3	1	Water	NA	NA	20.3	10.02	05/01/2025	09:33
4	ICV	1	Water	NA	NA	20.3	7.02	05/01/2025	09:35
5	CCV1	1	Water	NA	NA	20.2	2.01	05/01/2025	09:40
6	Q1925-01	1	Solid	20.02	20	24.4	9.42	05/01/2025	10:00
7	Q1925-01DUP	1	Solid	20.03	20	24.5	9.43	05/01/2025	10:01
8	Q1925-02	1	Solid	20.02	20	24.4	8.79	05/01/2025	10:20
9	Q1925-03	1	Solid	20.03	20	24.0	9.42	05/01/2025	10:30
10	CCV2	1	Water	NA	NA	20.2	12.02	05/01/2025	10:35

WORKLIST(Hardcopy Internal Chain)

135620

WorkList Name : PH S Q1925

WorkList ID : 189245

Department : Wet-Chemistry

Date : 05-01-2025 07:55:58

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1925-01	AUD-25-0068	Solid	pH	Cool 4 deg C	PSEG03	L31	04/30/2025	9045D
Q1925-02	AUD-25-0069	Solid	pH	Cool 4 deg C	PSEG03	L31	04/30/2025	9045D
Q1925-03	AUD-25-0070	Solid	pH	Cool 4 deg C	PSEG03	L31	04/30/2025	9045D

Date/Time 05/01/25 08:00
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: [Signature]

Date/Time 05/01/25 12:30
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: [Signature]