

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Iwona

Parameter: Flash Point

Supervisor Review By: jignesh

Run Number: LB135628

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3193

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.1	05/01/2025	09:45
2	Q1913-02		1	12	100.00	>212.0	>212.0	05/01/2025	10:15
3	Q1913-02DUP		1	13	100.00	>212.0	>212.0	05/01/2025	10:45
4	Q1913-04		1	14	100.00	>212.0	>212.0	05/01/2025	11:15
5	Q1919-05		1	12	100.00	>212.0	>212.0	05/01/2025	11:45
6	Q1919-08		1	13	100.00	>212.0	>212.0	05/01/2025	12:15
7	Q1925-04		1	12	100.00	>212.0	>212.0	05/01/2025	12:45
8	Q1925-05		1	13	100.00	>212.0	>212.0	05/01/2025	13:15
9	Q1926-01		1	12	100.00	>212.0	>212.0	05/01/2025	13:45

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

15135628

WORKLIST(Hardcopy Internal Chain)

WorkList Name : fp 5-1

WorkList ID : 189265

Department : Wet-Chemistry

Date : 05-01-2025 09:18:12

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1913-02	WC-12-A-202504	Water	Flash Point	Cool 4 deg C	AECO02	L21	04/29/2025	1010B
Q1913-04	WC-13-A-202504	Water	Flash Point	Cool 4 deg C	AECO02	L21	04/29/2025	1010B
Q1919-05	0423	Water	Flash Point	Cool 4 deg C	PSEG03	L41	04/30/2025	1010B
Q1919-08	0428	Water	Flash Point	Cool 4 deg C	PSEG03	L41	04/30/2025	1010B
Q1925-04	AUD-25-0072	Water	Flash Point	Cool 4 deg C	PSEG03	L31	04/30/2025	1010B
Q1925-05	AUD-25-0073	Water	Flash Point	Cool 4 deg C	PSEG03	L31	04/30/2025	1010B
Q1926-01	TRE-1104	Water	Flash Point	Cool 4 deg C	PSEG03	L31	04/30/2025	1010B

Date/Time

05/01/25 09:30

Raw Sample Received by:

12/20/

Raw Sample Relinquished by:

12/20/

Date/Time

05/01/25 14:30

Raw Sample Received by:

12/20/

Raw Sample Relinquished by:

12/20/