

Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB135640

Analyst By : jignesh

Supervisor Review By : rubina

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3178
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3191
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3161
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	05/02/2025	08:20
2	CAL2	1	Water	NA	NA	20.2	7.00	05/02/2025	08:21
3	CAL3	1	Water	NA	NA	20.3	10.02	05/02/2025	08:22
4	ICV	1	Water	NA	NA	20.3	7.00	05/02/2025	08:25
5	CCV1	1	Water	NA	NA	20.3	2.01	05/02/2025	08:30
6	Q1931-01	1	Water	NA	NA	23.1	6.64	05/02/2025	08:40
7	Q1931-01DUP	1	Water	NA	NA	23.2	6.65	05/02/2025	08:41
8	CCV2	1	Water	NA	NA	20.2	12.00	05/02/2025	08:45

WORKLIST(Hardcopy Internal Chain)

JB 132640

WorkList Name : ph w q1931

WorkList ID : 189274

Department : Wet-Chemistry

Date : 05-02-2025 08:04:05

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1931-01	NWB-2119	Water	pH	Cool 4 deg C	PSEG03	L51	05/01/2025	9040C

Date/Time 05/02/25 08:15

Raw Sample Received by: JRCWC

Raw Sample Relinquished by: chsn

Date/Time

Raw Sample Received by: JRCWC

Raw Sample Relinquished by: JRCWC