

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB135647

Slope : 99.2

BalanceID: WC SC-7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3178
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3191
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3161
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	05/02/2025	15:25
2	CAL2	1	Water	NA	NA	20.2	7.00	05/02/2025	15:26
3	CAL3	1	Water	NA	NA	20.2	10.02	05/02/2025	15:30
4	ICV	1	Water	NA	NA	20.3	7.01	05/02/2025	15:35
5	CCV1	1	Water	NA	NA	20.2	2.01	05/02/2025	15:40
6	Q1944-01	1	Solid	5.00	10	23.7	3.88	05/02/2025	15:50
7	CCV2	1	Water	NA	NA	20.2	12.02	05/02/2025	15:51

135647

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH S Q1944

WorkList ID : 189291

Department : Wet-Chemistry

Date : 05-02-2025 15:15:27

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1944-01	MELTED-RUBBER-SAMPLE	Solid	pH	Cool 4 deg C	METE01	L41	05/02/2025	9045D

Date/Time 05/02/25 15:20

Raw Sample Received by: YD Gull

Raw Sample Relinquished by: SL29 met-di's

Date/Time 05/02/25 17:30

Raw Sample Received by: SL29 met-di's

Raw Sample Relinquished by: SL29 met-di's