

Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB135648

Slope : 99.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3178
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3191
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3161
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	05/02/2025	15:25
2	CAL2	1	Water	NA	NA	20.2	7.00	05/02/2025	15:26
3	CAL3	1	Water	NA	NA	20.2	10.02	05/02/2025	15:30
4	ICV	1	Water	NA	NA	20.3	7.01	05/02/2025	15:35
5	CCV1	1	Water	NA	NA	20.2	2.01	05/02/2025	15:40
6	Q1944-02	1	Water	NA	NA	23.6	7.65	05/02/2025	16:00
7	Q1944-02DUP	1	Water	NA	NA	23.7	7.66	05/02/2025	16:02
8	Q1944-03	1	Water	NA	NA	24.1	8.99	05/02/2025	16:15
9	CCV2	1	Water	NA	NA	20.3	12.02	05/02/2025	16:20

WORKLIST(Hardcopy Internal Chain)

UB135648

WorkList Name : PH W Q1944 WorkList ID : 189292 Department : Wet-Chemistry Date : 05-02-2025 15:16:14

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1944-02	B CITY-WATER	Water	pH	Cool 4 deg C	METE01	L41	05/01/2025	9040C
Q1944-03	B CHILLER-WATER	Water	pH	Cool 4 deg C	METE01	L41	05/01/2025	9040C

Date/Time 05/02/25 15:20
Raw Sample Received by: SPULLY
Raw Sample Relinquished by: SPULLY

Date/Time 05/02/25 17:30
Raw Sample Received by: SPULLY
Raw Sample Relinquished by: SPULLY