

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB135675

Slope : 98.6

BalanceID: WC SC-7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3178
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3191
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3161
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	05/05/2025	15:00
2	CAL2	1	Water	NA	NA	20.2	7.00	05/05/2025	15:01
3	CAL3	1	Water	NA	NA	20.2	10.02	05/05/2025	15:05
4	ICV	1	Water	NA	NA	20.3	7.01	05/05/2025	15:10
5	CCV1	1	Water	NA	NA	20.2	2.01	05/05/2025	15:11
6	Q1956-05	1	Solid	20.02	20	21.2	7.52	05/05/2025	15:20
7	Q1956-05DUP	1	Solid	20.03	20	21.3	7.53	05/05/2025	15:21
8	Q1960-03	1	Solid	20.04	20	22.2	7.05	05/05/2025	15:30
9	CCV2	1	Water	NA	NA	20.3	12.02	05/05/2025	15:35

WORKLIST(Hardcopy Internal Chain)

UB 135675

WorkList Name : corrosivity q1956 WorkList ID : 189321 Department : Wet-Chemistry Date : 05-05-2025 11:15:30

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1956-05	COMP1	Solid	Corrosivity	Cool 4 deg C	CAMP02	L31	05/02/2025	9045D
Q1960-03	72-11933	Solid	Corrosivity	Cool 4 deg C	PSEG03	L51	05/05/2025	9045D

Date/Time 05/05/25 14:30
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: [Signature]

Date/Time 05/05/25 17:30
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: [Signature]