

LB135678


7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

45684



A= 0.0000e+000

B= 2.1478e-004

C= -3.6117e-002

Rho= 0.9992313

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.064	-0.064	-129	0.000	-129					0
0.2	0.200	0.169	-0.031	953	0.0 %	953					-16
2.5	2.500	2.550	0.050	12040	0.0 %	12040					2
5.0	5.000	5.010	0.010	23493	0.0 %	23493					0
7.5	7.500	7.760	0.260	36299	0.0 %	36299					3
10.0	10.000	9.776	-0.224	45684	0.0 %	45684					-2

LB135678 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 06 May 2025 09:50:00

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-129	-	0.0000	7471B	PPB	06 May 2025 10:00:06	S	Std
0.2 - 1	50.2	953	-	0.2000	7471B	PPB	06 May 2025 10:02:23	S	Std
2.5 - 1	52.5	12040	-	2.5000	7471B	PPB	06 May 2025 10:04:40	S	Std
5.0 - 1	55	23493	-	5.0000	7471B	PPB	06 May 2025 10:06:57	S	Std
7.5 - 1	57.5	36299	-	7.5000	7471B	PPB	06 May 2025 10:14:37	S	Std
10.0 - 1	510	45684	-	10.0000	7471B	PPB	06 May 2025 10:19:57	S	Std
ICV117 - 1	ICV117	19065	4.0586	-7471B	PPB	06 May 2025 10:25:28	U	SMPL	
ICB117 - 1	ICB117	-290	-0.0984	-7471B	PPB	06 May 2025 10:27:43	U	SMPL	
CCV82 - 1	CCV82	23733	5.0612	-7471B	PPB	06 May 2025 10:30:00	U	SMPL	
CCB82 - 1	CCB82	4	-0.0353	-7471B	PPB	06 May 2025 10:35:04	U	SMPL	
CRA - 1	CRA	1009	0.1806	-7471B	PPB	06 May 2025 10:37:19	U	SMPL	
HighStd - 1	HighStd	45492	9.7345	-7471B	PPB	06 May 2025 10:39:33	U	SMPL	
ChkStd - 1	ChkStd	33597	7.1797	-7471B	PPB	06 May 2025 10:41:49	U	SMPL	
PB167875BL - 1	PBS	-294	-0.0993	-7471B	PPB	06 May 2025 10:44:10	U	SMPL	
PB167875BS - 1	LCSS	18674	3.9746	-7471B	PPB	06 May 2025 10:49:16	U	SMPL	
Q1936-01 - 1	SMALL-PILE-A	281	0.0242	-7471B	PPB	06 May 2025 10:51:34	U	SMPL	
Q1936-01DUP - 1	SMALL-PILE-ADUP	406	0.0511	-7471B	PPB	06 May 2025 10:53:53	U	SMPL	
Q1936-01MS - 1	SMALL-PILE-AMS	18426	3.9214	-7471B	PPB	06 May 2025 10:56:10	U	SMPL	
Q1936-01MSD - 1	SMALL-PILE-AMSD	18235	3.8803	-7471B	PPB	06 May 2025 10:58:28	U	SMPL	
Q1936-03 - 1	SMALL-PILE-B	881	0.1531	-7471B	PPB	06 May 2025 11:00:46	U	SMPL	
CCV83 - 1	CCV83	23523	5.0161	-7471B	PPB	06 May 2025 11:03:04	U	SMPL	
CCB83 - 1	CCB83	0	-0.0361	-7471B	PPB	06 May 2025 11:05:21	U	SMPL	
Q1936-05 - 1	SMALL-PILE-C	824	0.1409	-7471B	PPB	06 May 2025 11:07:40	U	SMPL	
Q1936-07 - 1	SMALL-PILE-D	262	0.0202	-7471B	PPB	06 May 2025 11:09:56	U	SMPL	
Q1937-01 - 1	LARGE-PILE-A	695	0.1132	-7471B	PPB	06 May 2025 11:12:12	U	SMPL	
Q1937-03 - 1	LARGE-PILE-B	293	0.0268	-7471B	PPB	06 May 2025 11:14:28	U	SMPL	
Q1937-05 - 1	LARGE-PILE-C	289	0.0260	-7471B	PPB	06 May 2025 11:16:43	U	SMPL	
Q1937-07 - 1	LARGE-PILE-D	667	0.1071	-7471B	PPB	06 May 2025 11:19:00	U	SMPL	
Q1937-09 - 1	LARGE-PILE-E	2190	0.4342	-7471B	PPB	06 May 2025 11:21:16	U	SMPL	
Q1937-11 - 1	LARGE-PILE-F	2370	0.4729	-7471B	PPB	06 May 2025 11:23:33	U	SMPL	
Q1938-01 - 1	LOWER-WALL-PILE-A	3056	0.6202	-7471B	PPB	06 May 2025 11:25:51	U	SMPL	
Q1938-03 - 1	LOWER-WALL-PILE-B	2647	0.5324	-7471B	PPB	06 May 2025 11:28:07	U	SMPL	
CCV84 - 1	CCV84	24342	5.1920	-7471B	PPB	06 May 2025 11:30:24	U	SMPL	
CCB84 - 1	CCB84	-9	-0.0380	-7471B	PPB	06 May 2025 11:32:41	U	SMPL	
Q1938-05 - 1	LOWER-WALL-PILE-C	4398	0.9085	-7471B	PPB	06 May 2025 11:35:00	U	SMPL	
Q1938-07 - 1	LOWER-WALL-PILE-D	2161	0.4280	-7471B	PPB	06 May 2025 11:37:15	U	SMPL	
Q1946-01 - 1	BB-CONCRETE	205	0.0079	-7471B	PPB	06 May 2025 11:39:33	U	SMPL	
Q1947-01 - 1	MH-P	1599	0.3073	-7471B	PPB	06 May 2025 11:41:48	U	SMPL	
Q1947-05 - 1	MH-O	214	0.0098	-7471B	PPB	06 May 2025 11:44:04	U	SMPL	
Q1951-01 - 1	MH-KK	1442	0.2736	-7471B	PPB	06 May 2025 11:46:20	U	SMPL	
Q1952-01 - 1	CLEAN-FILL	2852	0.5764	-7471B	PPB	06 May 2025 11:48:37	U	SMPL	
Q1955-01 - 1	LAW-25-0066	1271	0.2369	-7471B	PPB	06 May 2025 11:50:54	U	SMPL	
Q1936-01LX5 - 1		61	-0.0230	-7471B	PPB	06 May 2025 11:53:13	U	SMPL	
Q1936-01A - 1		18562	3.9506	-7471B	PPB	06 May 2025 11:55:29	U	SMPL	
CCV85 - 1	CCV85	23429	4.9959	-7471B	PPB	06 May 2025 11:57:45	U	SMPL	
CCB85 - 1	CCB85	10	-0.0340	-7471B	PPB	06 May 2025 12:00:04	U	SMPL	