

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB135684

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: Flash Point

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3193

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	9	28.00	82.4	82.1	05/06/2025	12:50
2	Q1931-01		1	14	100.00	>212.0	>212.0	05/06/2025	13:20
3	Q1931-01DUP		1	14	100.00	>212.0	>212.0	05/06/2025	13:50

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

16135684

WORKLIST(Hardcopy Internal Chain)

WorkList Name : FP-5-6

WorkList ID : 189343

Department : Wet-Chemistry

Date : 05-06-2025 08:49:01

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1931-01	NWB-2119	Water	Flash Point	Cool 4 deg C	PSEG03	L51	05/01/2025	1010B

Date/Time 05/06/2025 12:40
Raw Sample Received by: RM CWL
Raw Sample Relinquished by: AF CWL

Date/Time 05/06/2025 14:30
Raw Sample Received by: AF CWL
Raw Sample Relinquished by: RM CWL