

Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB135686

Analyst By : jignesh

Supervisor Review By : Iwona

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3178
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3191
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3161
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	05/06/2025	15:10
2	CAL2	1	Water	NA	NA	20.2	7.00	05/06/2025	15:11
3	CAL3	1	Water	NA	NA	20.3	10.02	05/06/2025	15:15
4	ICV	1	Water	NA	NA	20.2	7.01	05/06/2025	15:17
5	CCV1	1	Water	NA	NA	20.2	2.01	05/06/2025	15:20
6	Q1965-03	1	Water	NA	NA	20.7	7.64	05/06/2025	15:30
7	Q1965-03DUP	1	Water	NA	NA	20.8	7.65	05/06/2025	15:31
8	Q1965-04	1	Water	NA	NA	20.1	7.60	05/06/2025	15:35
9	CCV2	1	Water	NA	NA	20.2	12.02	05/06/2025	15:40

WB 135686

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph q1965 WorkList ID : 189349 Department : Wet-Chemistry Date : 05-06-2025 14:47:40

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1965-03	A 0502	Water	pH	Cool 4 deg C	PSEG03	L41	05/06/2025	9040C
Q1965-04	A 0505	Water	pH	Cool 4 deg C	PSEG03	L41	05/06/2025	9040C

Date/Time 05/06/25 15:00
Raw Sample Received by: JP (57C)
Raw Sample Relinquished by: JP (57C)

Date/Time 05/06/25 17:00
Raw Sample Received by: JP (57C)
Raw Sample Relinquished by: JP (57C)