

LB135690


7470A

INSTRUMENT ID: CV2

Linear

 μ Abs.:

4707E



A= 0.0000e+000

B= 2.1182e-004

C= 3.6054e-002

Rho= 0.9996849

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%
0.0	0.000	0.039	0.039	15	0.000	15	0				0/00
0.2	0.200	0.226	0.026	898	0.0 %	898					13
2.5	2.500	2.345	-0.155	10898	0.0 %	10898					-6
5.0	5.000	5.145	0.145	24119	0.0 %	24119					3
7.5	7.500	7.437	-0.063	34939	0.0 %	34939					-1
10.0	10.000	10.008	0.008	47076	0.0 %	47076					0

LB135690 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 07 May 2025 08:24:03

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	15	-	0.0000	7470A	PPB	07 May 2025 09:05:42	S	Std
0.2 - 1	50.2	898	-	0.2000	7470A	PPB	07 May 2025 09:07:59	S	Std
2.5 - 1	52.5	10898	-	2.5000	7470A	PPB	07 May 2025 09:10:16	S	Std
5.0 - 1	55	24119	-	5.0000	7470A	PPB	07 May 2025 09:15:15	S	Std
7.5 - 1	57.5	34939	-	7.5000	7470A	PPB	07 May 2025 09:17:31	S	Std
10.0 - 1	57.6	47076	-	10.0000	7470A	PPB	07 May 2025 09:19:49	S	Std
ICV01 - 1	ICV01	17190	3.6773	-	7470A	PPB	07 May 2025 09:22:52	U	SMPL
ICB01 - 1	ICB01	-288	-0.0250	-	7470A	PPB	07 May 2025 09:25:06	U	SMPL
CCV86 - 1	CCV86	23582	5.0313	-	7470A	PPB	07 May 2025 09:27:23	U	SMPL
CCB86 - 1	CCB86	-330	-0.0338	-	7470A	PPB	07 May 2025 09:29:38	U	SMPL
CRA - 1	CRA	771	0.1994	-	7470A	PPB	07 May 2025 09:31:55	U	SMPL
HighStd - 1	HighStd	47700	10.1401	-	7470A	PPB	07 May 2025 09:34:10	U	SMPL
ChkStd - 1	ChkStd	28617	6.0978	-	7470A	PPB	07 May 2025 09:41:51	U	SMPL
PB167883BS - 1	LCSW	19357	4.1363	-	7470A	PPB	07 May 2025 09:46:25	U	SMPL
Q1947-04 - 1	MH-P	-235	-0.0137	-	7470A	PPB	07 May 2025 09:48:42	U	SMPL
Q1947-08 - 1	MH-O	-21	0.0316	-	7470A	PPB	07 May 2025 09:51:00	U	SMPL
Q1951-04 - 1	MH-KK	-11	0.0337	-	7470A	PPB	07 May 2025 09:53:17	U	SMPL
Q1956-05 - 1	COMP1	-2	0.0356	-	7470A	PPB	07 May 2025 09:55:34	U	SMPL
Q1959-01 - 1	SOIL-PILE	-57	0.0240	-	7470A	PPB	07 May 2025 09:57:49	U	SMPL
CCV87 - 1	CCV87	23476	5.0089	-	7470A	PPB	07 May 2025 10:00:05	U	SMPL
CCB87 - 1	CCB87	-317	-0.0311	-	7470A	PPB	07 May 2025 10:02:21	U	SMPL
Q1960-03 - 1	72-11933	-19	0.0320	-	7470A	PPB	07 May 2025 10:04:39	U	SMPL
Q1960-03DUP - 1	72-11933DUP	-19	0.0320	-	7470A	PPB	07 May 2025 10:06:54	U	SMPL
Q1960-03MS - 1	72-11933MS	19331	4.1308	-	7470A	PPB	07 May 2025 10:09:10	U	SMPL
Q1960-03MSD - 1	72-11933MSD	17835	3.8139	-	7470A	PPB	07 May 2025 10:11:25	U	SMPL
PB167851TB - 1	PB167851TB	-207	-0.0078	-	7470A	PPB	07 May 2025 10:13:43	U	SMPL
PB167883BL - 1	PBW	2	0.0365	-	7470A	PPB	07 May 2025 10:16:01	U	SMPL
Q1960-03LX5 - 1		-58	0.0238	-	7470A	PPB	07 May 2025 10:18:17	U	SMPL
Q1960-03A - 1		18496	3.9540	-	7470A	PPB	07 May 2025 10:20:34	U	SMPL
CCV88 - 1	CCV88	23493	5.0125	-	7470A	PPB	07 May 2025 10:22:51	U	SMPL
CCB88 - 1	CCB88	-291	-0.0256	-	7470A	PPB	07 May 2025 10:25:08	U	SMPL