


LB135692

7471B

INSTRUMENT ID: CW

Linear



A= 0.0000e+000

B= 1.6369e-004

C= -2.1812e-002

Rho= 0.9997684

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	σ/1
0.0	0.000	-0.013	-0.013	56	0.000	56					-
0.2	0.200	0.174	-0.026	1195	0.0 %	1195					-13
2.5	2.500	2.468	-0.032	15208	0.0 %	15208					-1
5.0	5.000	5.070	0.070	31107	0.0 %	31107					1
7.5	7.500	7.625	0.125	46714	0.0 %	46714					2
10.0	10.000	9.876	-0.124	60465	0.0 %	60465					-1

LB13566692 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 07 May 2025 12:35:32

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	56	-	0.0000	7471B	PPB	07 May 2025 12:39:50	S	Std
0.2 - 1	50.2	1195	-	0.2000	7471B	PPB	07 May 2025 12:42:07	S	Std
2.5 - 1	52.5	15208	-	2.5000	7471B	PPB	07 May 2025 12:44:24	S	Std
5.0 - 1	55	31107	-	5.0000	7471B	PPB	07 May 2025 12:46:41	S	Std
7.5 - 1	57.5	46714	-	7.5000	7471B	PPB	07 May 2025 12:49:00	S	Std
10.0 - 1	57.5	64249	-	10.0000	7471B	PPB	07 May 2025 12:51:20	S	Std
10.0 - 1	56.6	60465	-	10.0000	7471B	PPB	07 May 2025 12:54:04	S	Std
ICV02 - 1	ICV02	23674	3.8535	-7471B	PPB	PPB	07 May 2025 12:57:02	U	SMPL
ICB02 - 1	ICB02	-425	-0.0914	-7471B	PPB	PPB	07 May 2025 12:59:19	U	SMPL
CCV89 - 1	CCV89	31269	5.0967	-7471B	PPB	PPB	07 May 2025 13:01:36	U	SMPL
CCB89 - 1	CCB89	-255	-0.0636	-7471B	PPB	PPB	07 May 2025 13:03:52	U	SMPL
CRA - 1	CRA	1160	0.1681	-7471B	PPB	PPB	07 May 2025 13:06:11	U	SMPL
HighStd - 1	HighStd	59486	9.7157	-7471B	PPB	PPB	07 May 2025 13:08:26	U	SMPL
ChkStd - 1	ChkStd	44079	7.1937	-7471B	PPB	PPB	07 May 2025 13:10:41	U	SMPL
PB167884BL - 1	PBS	-7	-0.0230	-7471B	PPB	PPB	07 May 2025 13:17:01	U	SMPL
Q1960-02 - 1	72-11933	10731	1.7348	-7471B	PPB	PPB	07 May 2025 13:21:33	U	SMPL
Q1960-02DUP - 1	72-11933DUP	18762	3.0494	-7471B	PPB	PPB	07 May 2025 13:23:49	U	SMPL
Q1960-02MS - 1	72-11933MS	0	-0.0218	-7471B	PPB	PPB	07 May 2025 13:26:10	U	SMPL
Q1960-02 - 1	72-11933	13403	2.1722	-7471B	PPB	PPB	07 May 2025 13:26:39	U	SMPL
Q1960-02DUP - 1	72-11933DUP	13024	2.1101	-7471B	PPB	PPB	07 May 2025 13:28:56	U	SMPL
Q1960-02MS - 1	72-11933MS	38981	6.3592	-7471B	PPB	PPB	07 May 2025 13:33:57	U	SMPL
Q1960-02MSD - 1	72-11933MSD	36380	5.9334	-7471B	PPB	PPB	07 May 2025 13:36:14	U	SMPL
Q1961-01 - 1	EO-01-050525	2344	0.3619	-7471B	PPB	PPB	07 May 2025 13:38:35	U	SMPL
PB167884BS - 1	LCSS	23179	3.7725	-7471B	PPB	PPB	07 May 2025 13:41:22	U	SMPL
CCV90 - 1	CCV90	30369	4.9494	-7471B	PPB	PPB	07 May 2025 13:43:59	U	SMPL
CCB90 - 1	CCB90	-296	-0.0703	-7471B	PPB	PPB	07 May 2025 13:46:16	U	SMPL
Q1962-01 - 1	HD-01-050525	3274	0.5141	-7471B	PPB	PPB	07 May 2025 13:48:34	U	SMPL
Q1964-01 - 1	MH-LL	4893	0.7791	-7471B	PPB	PPB	07 May 2025 13:50:49	U	SMPL
Q1965-01 - 1	50623	-119	-0.0413	-7471B	PPB	PPB	07 May 2025 13:53:07	U	SMPL
Q1966-01 - 1	CONCRETE-1	916	0.1281	-7471B	PPB	PPB	07 May 2025 13:55:25	U	SMPL
Q1966-02 - 1	CONCRETE-2	3484	0.5485	-7471B	PPB	PPB	07 May 2025 13:57:41	U	SMPL
Q1966-03 - 1	SILICA-GEL	164	0.0050	-7471B	PPB	PPB	07 May 2025 13:59:58	U	SMPL
Q1968-01 - 1	MH-N	455	0.0527	-7471B	PPB	PPB	07 May 2025 14:02:15	U	SMPL
Q1968-05 - 1	MH-M	683	0.0900	-7471B	PPB	PPB	07 May 2025 14:04:32	U	SMPL
Q1960-02LX5 - 1		2969	0.4642	-7471B	PPB	PPB	07 May 2025 14:06:49	U	SMPL
Q1960-02A - 1		34514	5.6279	-7471B	PPB	PPB	07 May 2025 14:09:05	U	SMPL
CCV91 - 1	CCV91	30271	4.9334	-7471B	PPB	PPB	07 May 2025 14:11:22	U	SMPL
CCB91 - 1	CCB91	-388	-0.0853	-7471B	PPB	PPB	07 May 2025 14:13:45	U	SMPL