

Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB135700

Analyst By : jignesh

Supervisor Review By : rubina

Slope : 99.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3178
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3191
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3161
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	05/08/2025	08:15
2	CAL2	1	Water	NA	NA	20.2	7.00	05/08/2025	08:16
3	CAL3	1	Water	NA	NA	20.2	10.02	05/08/2025	08:19
4	ICV	1	Water	NA	NA	20.2	7.00	05/08/2025	08:22
5	CCV1	1	Water	NA	NA	20.1	2.01	05/08/2025	08:25
6	Q1980-01	1	Water	NA	NA	22.1	7.62	05/08/2025	08:35
7	Q1980-01DUP	1	Water	NA	NA	22.2	7.65	05/08/2025	08:37
8	Q1980-02	1	Water	NA	NA	22.2	6.67	05/08/2025	08:45
9	Q1980-02DUP	1	Water	NA	NA	22.3	6.70	05/08/2025	08:49
10	Q1980-03	1	Water	NA	NA	22.3	6.54	05/08/2025	09:00
11	Q1980-03DUP	1	Water	NA	NA	22.4	6.55	05/08/2025	09:05
12	CCV2	1	Water	NA	NA	20.3	12.02	05/08/2025	09:11

WORKLIST(Hardcopy Internal Chain)

VB 135700

WorkList Name : ph w q1980

WorkList ID : 189375

Department : Wet-Chemistry

Date : 05-08-2025 07:53:36

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1980-01	A 3834	Water	pH	Cool 4 deg C	PSEG03	L61	05/07/2025	9040C
Q1980-02	C 3832	Water	pH	Cool 4 deg C	PSEG03	L61	05/07/2025	9040C
Q1980-03	C 3835	Water	pH	Cool 4 deg C	PSEG03	L61	05/07/2025	9040C

Date/Time 05/08/25 08:10

Raw Sample Received by: 160201

Raw Sample Relinquished by: SA SM

Date/Time

05/08/25

Raw Sample Received by:

Raw Sample Relinquished by:

11:00

SA SM

SA SM