

LB135701


7471B

INSTRUMENT ID: CV2

Linear

 μ Abs.:

54432



A= 0.0000e+000

B= 1.8610e-004

C= -5.9859e-002

Rho= 0.9998349

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	Q/A
0.0	0.000	-0.005	-0.005	295	0.000	295					-2
0.2	0.200	0.197	-0.003	1381	0.0 %	1381					1
2.5	2.500	2.528	0.028	13907	0.0 %	13907					1
5.0	5.000	5.047	0.047	27444	0.0 %	27444					1
7.5	7.500	7.362	-0.138	39882	0.0 %	39882					-2
10.0	10.000	10.070	0.070	54432	0.0 %	54432					1

LB135701 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 08 May 2025 09:45:50

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	295	-	0.0000	7471B	PPB	08 May 2025 09:46:01	S	Std
0.2 - 1	50.2	1381	-	0.2000	7471B	PPB	08 May 2025 09:48:18	S	Std
2.5 - 1	52.5	13907	-	2.5000	7471B	PPB	08 May 2025 09:50:35	S	Std
5.0 - 1	55	27444	-	5.0000	7471B	PPB	08 May 2025 09:52:52	S	Std
7.5 - 1	57.5	39882	-	7.5000	7471B	PPB	08 May 2025 09:55:11	S	Std
10.0 - 1	510	54432	-	10.0000	7471B	PPB	08 May 2025 09:57:29	S	Std
ICV03 - 1	ICV03	19754	3.6164	-	7471B	PPB	08 May 2025 10:00:31	U	SMPL
ICB03 - 1	ICB03	22	-0.0558	-	7471B	PPB	08 May 2025 10:02:47	U	SMPL
CCV01 - 1	CCV01	27322	5.0248	-	7471B	PPB	08 May 2025 10:05:05	U	SMPL
CCB01 - 1	CCB01	-24	-0.0643	-	7471B	PPB	08 May 2025 10:07:20	U	SMPL
CRA - 1	CRA	1285	0.1793	-	7471B	PPB	08 May 2025 10:09:38	U	SMPL
HighStd - 1	HighStd	53715	9.9365	-	7471B	PPB	08 May 2025 10:11:53	U	SMPL
ChkStd - 1	ChkStd	37556	6.9293	-	7471B	PPB	08 May 2025 10:14:09	U	SMPL
PB167911BL - 1	PBS	351	0.0055	-	7471B	PPB	08 May 2025 10:19:17	U	SMPL
PB167911BS - 1	LCSSSSSSS	19377	3.5462	-	7471B	PPB	08 May 2025 10:21:32	U	SMPL
Q1969-01 - 1	TR-06-05072025	965	0.1197	-	7471B	PPB	08 May 2025 10:23:49	U	SMPL
Q1969-01DUP - 1	TR-06-05072025DUP	836	0.0957	-	7471B	PPB	08 May 2025 10:31:54	U	SMPL
Q1969-01MS - 1	TR-06-05072025MS	22189	4.0695	-	7471B	PPB	08 May 2025 10:34:11	U	SMPL
Q1969-01MSD - 1	TR-06-05072025MSD	20496	3.7544	-	7471B	PPB	08 May 2025 10:36:29	U	SMPL
Q1972-01 - 1	SUB-WC	1791	0.2734	-	7471B	PPB	08 May 2025 10:38:48	U	SMPL
CCV02 - 1	CCV02	26194	4.8148	-	7471B	PPB	08 May 2025 10:41:09	U	SMPL
CCB02 - 1	CCB02	128	-0.0360	-	7471B	PPB	08 May 2025 10:43:26	U	SMPL
Q1972-05 - 1	WC-B-10	647	0.0605	-	7471B	PPB	08 May 2025 10:45:45	U	SMPL
Q1972-09 - 1	WC-B-2	729	0.0758	-	7471B	PPB	08 May 2025 10:48:00	U	SMPL
Q1972-13 - 1	WC-B-7	1201	0.1636	-	7471B	PPB	08 May 2025 10:50:16	U	SMPL
Q1972-17 - 1	SUB-WC-2	1674	0.2517	-	7471B	PPB	08 May 2025 10:52:32	U	SMPL
Q1972-21 - 1	WC-B-1	2994	0.4973	-	7471B	PPB	08 May 2025 10:54:47	U	SMPL
Q1974-01 - 1	OILY-DEBRIS-COMP	70	-0.0468	-	7471B	PPB	08 May 2025 10:57:06	U	SMPL
Q1974-03 - 1	OILY-PADS	648	0.0607	-	7471B	PPB	08 May 2025 10:59:23	U	SMPL
Q1976-01 - 1	VNJ-207	3970	0.6790	-	7471B	PPB	08 May 2025 11:01:41	U	SMPL
Q1978-01 - 1	72-12016	3297	0.5537	-	7471B	PPB	08 May 2025 11:03:57	U	SMPL
CCV03 - 1	CCV03	26414	4.8558	-	7471B	PPB	08 May 2025 11:06:15	U	SMPL
CCB03 - 1	CCB03	84	-0.0442	-	7471B	PPB	08 May 2025 11:08:32	U	SMPL
Q1969-01LX5 - 1		121	-0.0373	-	7471B	PPB	08 May 2025 11:10:50	U	SMPL
Q1969-01A - 1		21151	3.8763	-	7471B	PPB	08 May 2025 11:13:06	U	SMPL
CCV04 - 1	CCV04	26024	4.7832	-	7471B	PPB	08 May 2025 11:15:21	U	SMPL
CCB04 - 1	CCB04	-60	-0.0710	-	7471B	PPB	08 May 2025 11:17:41	U	SMPL