

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB135709

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: Flash Point

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3193

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.4	05/09/2025	09:20
2	Q1980-01		1	15	100.00	>212.0	>212.0	05/09/2025	09:50
3	Q1980-01DUP		1	15	100.00	>212.0	>212.0	05/09/2025	10:20
4	Q1980-02		1	12	100.00	>212.0	>212.0	05/09/2025	10:50
5	Q1980-03		1	14	100.00	>212.0	>212.0	05/09/2025	11:20

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

06135709

WorkList Name : fp-05-08 WorkList ID : 199379 Department : Wet-Chemistry Date : 05-08-2025 08:22:23

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1980-01	3834	Water	Flash Point	Cool 4 deg C	PSEG03	L61	05/07/2025	1010B
Q1980-02	3832	Water	Flash Point	Cool 4 deg C	PSEG03	L61	05/07/2025	1010B
Q1980-03	3835	Water	Flash Point	Cool 4 deg C	PSEG03	L61	05/07/2025	1010B

Date/Time 05/09/2025 09:10
 Raw Sample Received by: RM
 Raw Sample Relinquished by: RM

Date/Time 05/09/2025 12:06
 Raw Sample Received by: RM
 Raw Sample Relinquished by: RM