

Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB135731

Analyst By : jignesh

Supervisor Review By : Iwona

Slope : 98.4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3178
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3191
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3161
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.05].

True Value of CCV1 = 2.00 Control Limits[+/- 0.05].

True Value of CCV2 = 12.00 Control Limits[+/- 0.05].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	05/12/2025	08:30
2	CAL2	1	Water	NA	NA	20.3	7.01	05/12/2025	08:32
3	CAL3	1	Water	NA	NA	20.3	10.02	05/12/2025	08:35
4	ICV	1	Water	NA	NA	20.2	7.00	05/12/2025	08:44
5	CCV1	1	Water	NA	NA	20.3	2.01	05/12/2025	08:47
6	Q2005-01	1	Water	NA	NA	20.7	7.07	05/12/2025	08:55
7	Q2005-01DUP	1	Water	NA	NA	20.8	7.08	05/12/2025	08:56
8	CCV2	1	Water	NA	NA	20.3	12.02	05/12/2025	09:00

WORKLIST(Hardcopy Internal Chain)

VB 135f31

WorkList Name : PH W Q2005 WorkList ID : 189430 Department : Wet-Chemistry Date : 05-12-2025 08:10:57

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q2005-01- <u>✓</u>	252806	Water	pH	Cool 4 deg C	PSEG03	L41	05/09/2025	9040C

Date/Time 05/12/25 08:20
Raw Sample Received by: rf wci
Raw Sample Relinquished by: cf sm

Date/Time 05/12/25 12:30
Raw Sample Received by: cf sm
Raw Sample Relinquished by: rf wci