

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Iwona

Parameter: Flash Point

Supervisor Review By: jignesh

Run Number: LB135734

Ambient Barometric Pressure (mmHg): 771.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3193

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	27.00	80.6	79.9	05/12/2025	09:15
2	Q1988-01		1	14	100.00	>212.0	>212.0	05/12/2025	09:45
3	Q1988-01DUP		1	11	100.00	>212.0	>212.0	05/12/2025	10:15
4	Q1989-01		1	11	100.00	>212.0	>212.0	05/12/2025	10:45
5	Q1992-01		1	12	100.00	>212.0	>212.0	05/12/2025	11:15
6	Q2005-01		1	13	100.00	>212.0	>212.0	05/12/2025	11:45

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

16135734

WorkList Name : fp-w-05-12

WorkList ID : 189436

Department : Wet-Chemistry

Date : 05-12-2025 08:42:48

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1988-01	5725	Water	Flash Point	Cool 4 deg C	PSEG03	L41	05/08/2025	1010B
Q1989-01	50225	Water	Flash Point	Cool 4 deg C	PSEG03	L51	05/08/2025	1010B
Q1992-01	60453	Water	Flash Point	Cool 4 deg C	PSEG03	L41	05/08/2025	1010B
Q2005-01	252806	Water	Flash Point	Cool 4 deg C	PSEG03	L41	05/09/2025	1010B

Date/Time 05/12/25 09:05
 Raw Sample Received by: RLC
 Raw Sample Relinquished by: RLC

Date/Time 05/12/25 12:30
 Raw Sample Received by: RLC
 Raw Sample Relinquished by: RLC