

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB135737

Ambient Barometric Pressure (mmHg): 771.00

Thermometer ID: Flash Point

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3193

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	9	28.00	82.4	81.7	05/12/2025	11:25
2	Q1991-01		1	14	100.00	>212.0	>212.0	05/12/2025	11:55
3	Q1991-01DUP		1	14	100.00	>212.0	>212.0	05/12/2025	12:25
4	Q1991-03		1	13	100.00	>212.0	>212.0	05/12/2025	12:55

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06