

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB135750

Slope : 98.6

BalanceID: WC SC-7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3178
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3191
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3161
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.05].

True Value of CCV1 = 2.00 Control Limits[+/- 0.05].

True Value of CCV2 = 12.00 Control Limits[+/- 0.05].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	05/13/2025	09:00
2	CAL2	1	Water	NA	NA	20.2	7.00	05/13/2025	09:01
3	CAL3	1	Water	NA	NA	20.1	10.02	05/13/2025	09:03
4	ICV	1	Water	NA	NA	20.2	7.01	05/13/2025	09:05
5	CCV1	1	Water	NA	NA	20.3	2.01	05/13/2025	09:06
6	Q2017-04	1	Solid	20.02	20	22.4	7.12	05/13/2025	09:15
7	Q2017-04DUP	1	Solid	20.03	20	22.5	7.13	05/13/2025	09:16
8	Q2017-08	1	Solid	20.05	20	23.0	7.95	05/13/2025	09:30
9	CCV2	1	Water	NA	NA	20.2	12.02	05/13/2025	09:31

WORKLIST(Hardcopy Internal Chain)

135150

WorkList Name : corrosivity q2017

WorkList ID : 189472

Department : Wet-Chemistry

Date : 05-13-2025 08:30:11

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q2017-04	MH-I	Solid	Corrosivity	Cool 4 deg C	PSEG03	L41	05/12/2025	9045D
Q2017-08	MH-J	Solid	Corrosivity	Cool 4 deg C	PSEG03	L41	05/12/2025	9045D

Date/Time 05-13-25 08:40

Raw Sample Received by: [Signature]

Raw Sample Relinquished by: [Signature]

Date/Time 05-13-25

Raw Sample Received by: [Signature]

Raw Sample Relinquished by: [Signature]