

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Iwona

Parameter: Flash Point

Supervisor Review By: jignesh

Run Number: LB135765

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3193

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	29.00	84.2	83.9	05/14/2025	09:00
2	Q2008-01		1	12	100.00	>212.0	>212.0	05/14/2025	09:30
3	Q2008-01DUP		1	13	100.00	>212.0	>212.0	05/14/2025	10:00

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB135765

WorkList Name : FP-5-14

WorkList ID : 189506

Department : Wet-Chemistry

Date : 05-14-2025 08:13:39

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q2008-01	IDW-AQ-DRUM-633-05092025	Water	Flash Point	Cool 4 deg C	JACO05	L41	05/09/2025	1010B

Date/Time 05/14/25 08:50
 Raw Sample Received by: 12(50)
 Raw Sample Relinquished by: 286007

Date/Time 05/14/25 10:45
 Raw Sample Received by: 286007
 Raw Sample Relinquished by: 12(50)