

Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB135801

Analyst By : jignesh

Supervisor Review By : Iwona

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3178
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3191
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3161
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.05].

True Value of CCV1 = 2.00 Control Limits[+/- 0.05].

True Value of CCV2 = 12.00 Control Limits[+/- 0.05].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	05/16/2025	13:55
2	CAL2	1	Water	NA	NA	20.2	7.01	05/16/2025	13:56
3	CAL3	1	Water	NA	NA	20.3	10.02	05/16/2025	13:59
4	ICV	1	Water	NA	NA	20.3	7.01	05/16/2025	14:00
5	CCV1	1	Water	NA	NA	20.2	2.01	05/16/2025	14:02
6	Q2008-03	1	Water	NA	NA	20.7	2.03	05/16/2025	14:10
7	Q2008-03DUP	1	Water	NA	NA	21.8	2.04	05/16/2025	14:11
8	CCV2	1	Water	NA	NA	20.2	12.02	05/16/2025	14:15

135801

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH Q2008 WorkList ID : 189575 Department : Wet-Chemistry Date : 05-16-2025 13:29:07

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q2008-03	IDW-AQ-DRUM-633-05092025	Water	pH	Cool 4 deg C	JACO05	L41	05/09/2025	9040C

Date/Time 05/16/25 13:40
Raw Sample Received by: JBC
Raw Sample Relinquished by: JBC

Date/Time 05/16/25 16:10
Raw Sample Received by: JBC
Raw Sample Relinquished by: JBC