

Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB135814

Analyst By : jignesh

Supervisor Review By : Iwona

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3178
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3191
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3161
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.05].

True Value of CCV1 = 2.00 Control Limits[+/- 0.05].

True Value of CCV2 = 12.00 Control Limits[+/- 0.05].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	05/19/2025	08:25
2	CAL2	1	Water	NA	NA	20.2	7.00	05/19/2025	08:26
3	CAL3	1	Water	NA	NA	20.2	10.02	05/19/2025	08:29
4	ICV	1	Water	NA	NA	20.2	7.01	05/19/2025	08:33
5	CCV1	1	Water	NA	NA	20.2	2.01	05/19/2025	08:37
6	Q2065-01	1	Water	NA	NA	20.7	7.00	05/19/2025	08:45
7	Q2067-01	1	Water	NA	NA	20.7	7.02	05/19/2025	08:55
8	Q2067-01DUP	1	Water	NA	NA	20.8	7.03	05/19/2025	08:56
9	Q2068-03	1	Water	NA	NA	20.9	4.91	05/19/2025	09:10
10	CCV2	1	Water	NA	NA	20.2	12.02	05/19/2025	09:11

WORKLIST(Hardcopy Internal Chain)

VB35813

WorkList Name : ph w q2065

WorkList ID : 189585

Department : Wet-Chemistry

Date : 05-19-2025 08:10:20

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q2065-01 <i>Y</i>	VACTRUCK-728068	Water	pH	Cool 4 deg C	PSEG03	L41	05/16/2025	9040C
Q2067-01 <i>Y</i>	303-PPR-FRAC	Water	pH	Cool 4 deg C	PSEG03	L41	05/16/2025	9040C
Q2068-03 <i>Y</i>	LAW-25-0074	Water	pH	Cool 4 deg C	PSEG03	L31	05/16/2025	9040C

Date/Time 05-19-25 08:20
 Raw Sample Received by: JD w/c
 Raw Sample Relinquished by: CP

Date/Time 05-19-25 11:30
 Raw Sample Received by: CP
 Raw Sample Relinquished by: JD w/c