

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Iwona

Parameter: Flash Point

Supervisor Review By: jignesh

Run Number: LB135818

Ambient Barometric Pressure (mmHg): 755.00

Thermometer ID: Flash Point

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3193

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	100.00	>212.0	>212.0	05/19/2025	09:30
2	Q2044-01		1	4	6.00	42.8	43.1	05/19/2025	10:00
3	Q2044-01DUP		1	4	6.00	42.8	43.1	05/19/2025	10:30

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

6135818

WorkList Name : fp-05-19 WorkList ID : 189597 Department : Wet-Chemistry Date : 05-19-2025 08:19:38

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q2044-01	3858	Solid	Flash Point	Cool 4 deg C	PSEG03	L31	05/14/2025	1010B

Date/Time 05/19/2025 09:20
 Raw Sample Received by: 12(50)
 Raw Sample Relinquished by: HCO

Date/Time 05/19/2025 11:10
 Raw Sample Received by: HCO
 Raw Sample Relinquished by: 12(50)