

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Iwona

Parameter: Flash Point

Supervisor Review By: jignesh

Run Number: LB135829

Ambient Barometric Pressure (mmHg): 771.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3193

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	26.00	78.8	78.1	05/19/2025	13:30
2	Q2065-01		1	12	100.00	>212.0	>212.0	05/19/2025	14:00
3	Q2065-01DUP		1	12	100.00	>212.0	>212.0	05/19/2025	14:30
4	Q2067-01		1	13	100.00	>212.0	>212.0	05/19/2025	15:00
5	Q2068-03		1	13	100.00	>212.0	>212.0	05/19/2025	15:30

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

6135829

WorkList Name : FP-05-19*

WorkList ID : 189608

Department : Wet-Chemistry

Date : 05-19-2025 09:37:02

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q2065-01	VACTRUCK-728068	Water	Flash Point	Cool 4 deg C	PSEG03	L41	05/16/2025	1010B
Q2067-01	303-PPR-FRAC	Water	Flash Point	Cool 4 deg C	PSEG03	L41	05/16/2025	1010B
Q2068-03	LAW-25-0074	Water	Flash Point	Cool 4 deg C	PSEG03	L31	05/16/2025	1010B

Date/Time 05/19/25 13:10
 Raw Sample Received by: 12607
 Raw Sample Relinquished by: 12607

Date/Time 05/19/25 16:15
 Raw Sample Received by: 12607
 Raw Sample Relinquished by: 12607