

LB135832

7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

45590

A= 0.0000e+000

B= 2.1335e-004

C= 4.7703e-002

Rho= 0.9989599

Accept=Accepted

Max. Conc: —

10.000 PPB

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/1)
0.0	0.000	0.030	0.030	-85	0.000	-85	0				4
0.2	0.200	0.208	0.008	751	0.0 %	751					4
2.5	2.500	2.349	-0.151	10785	0.0 %	10785					6
5.0	5.000	5.033	0.033	23367	0.0 %	23367					1
7.5	7.500	7.807	0.307	36367	0.0 %	36367					4
10.0	10.000	9.774	-0.226	45590	0.0 %	45590					2

LB135832 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 20 May 2025 08:59:59

Sample ID	Extended ID	μ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	5.0	-85	-	0.0000	7470A	PPB	20 May 2025 09:10:11	S	Std
0.2 - 1	5.0	751	-	0.2000	7470A	PPB	20 May 2025 09:12:28	S	Std
2.5 - 1	5.0	10785	-	2.5000	7470A	PPB	20 May 2025 09:14:45	S	Std
5.0 - 1	5.0	23367	-	5.0000	7470A	PPB	20 May 2025 09:20:04	S	Std
7.5 - 1	5.0	36367	-	7.5000	7470A	PPB	20 May 2025 09:25:00	S	Std
10.0 - 1	5.0	45590	-	10.0000	7470A	PPB	20 May 2025 09:30:20	S	Std
ICV11 - 1	ICV11	19411	4.1890	-	7470A	PPB	20 May 2025 09:36:00	U	SMPL
ICB11 - 1	ICB11	-262	-0.0082	-	7470A	PPB	20 May 2025 09:38:15	U	SMPL
CCV40 - 1	CCV40	22963	4.9468	-	7470A	PPB	20 May 2025 09:43:19	U	SMPL
CCB40 - 1	CCB40	-360	-0.0291	-	7470A	PPB	20 May 2025 09:45:34	U	SMPL
CRA - 1	CRA	687	0.1943	-	7470A	PPB	20 May 2025 09:47:52	U	SMPL
HighStd - 1	HighStd	44737	9.5923	-	7470A	PPB	20 May 2025 09:50:07	U	SMPL
ChkStd - 1	ChkStd	29522	6.3462	-	7470A	PPB	20 May 2025 09:55:02	U	SMPL
PB168062BL - 1	PBW	-362	-0.0295	-	7470A	PPB	20 May 2025 09:57:17	U	SMPL
PB168062BS - 1	LCSW	16942	3.6623	-	7470A	PPB	20 May 2025 09:59:35	U	SMPL
Q2052-04 - 1	TP-B	-235	-0.0024	-	7470A	PPB	20 May 2025 10:01:51	U	SMPL
Q2053-01 - 1	SB-1	-73	0.0321	-	7470A	PPB	20 May 2025 10:04:10	U	SMPL
Q2053-02 - 1	SB-2	-15	0.0445	-	7470A	PPB	20 May 2025 10:06:26	U	SMPL
Q2053-03 - 1	SB-3	-1	0.0475	-	7470A	PPB	20 May 2025 10:08:44	U	SMPL
Q2053-04 - 1	SB-4	-32	0.0408	-	7470A	PPB	20 May 2025 10:11:00	U	SMPL
CCV41 - 1	CCV41	21695	4.6763	-	7470A	PPB	20 May 2025 10:13:15	U	SMPL
CCB41 - 1	CCB41	-377	-0.0327	-	7470A	PPB	20 May 2025 10:15:30	U	SMPL
Q2053-05 - 1	SB-5	29	0.0538	-	7470A	PPB	20 May 2025 10:17:49	U	SMPL
Q2053-06 - 1	SB-6	7	0.0492	-	7470A	PPB	20 May 2025 10:20:04	U	SMPL
Q2053-07 - 1	SB-7	2	0.0481	-	7470A	PPB	20 May 2025 10:22:19	U	SMPL
Q2053-08 - 1	SB-8	-10	0.0458	-	7470A	PPB	20 May 2025 10:24:35	U	SMPL
Q2053-09 - 1	SB-9	22	0.0524	-	7470A	PPB	20 May 2025 10:26:50	U	SMPL
Q2053-10 - 1	SB-10	-26	0.0422	-	7470A	PPB	20 May 2025 10:29:06	U	SMPL
Q2053-11 - 1	SB-11	-5	0.0466	-	7470A	PPB	20 May 2025 10:31:23	U	SMPL
Q2057-04 - 1	MH-L	-131	0.0198	-	7470A	PPB	20 May 2025 10:33:40	U	SMPL
Q2062-04 - 1	L1-WC-1	-96	0.0272	-	7470A	PPB	20 May 2025 10:35:57	U	SMPL
Q2062-08 - 1	L1-WC-2	-58	0.0353	-	7470A	PPB	20 May 2025 10:38:13	U	SMPL
CCV42 - 1	CCV42	22848	4.9223	-	7470A	PPB	20 May 2025 10:42:05	U	SMPL
CCB42 - 1	CCB42	-253	-0.0063	-	7470A	PPB	20 May 2025 10:44:20	U	SMPL
Q2062-12 - 1	L1-WC-3	-46	0.0379	-	7470A	PPB	20 May 2025 10:46:37	U	SMPL
Q2062-16 - 1	L1-WC-4	174	0.0849	-	7470A	PPB	20 May 2025 10:48:53	U	SMPL
Q2062-20 - 1	L1-WC-5	-26	0.0422	-	7470A	PPB	20 May 2025 10:51:08	U	SMPL
Q2062-24 - 1	L1-WC-6	-4	0.0468	-	7470A	PPB	20 May 2025 10:53:24	U	SMPL
Q2062-24DUP - 1	L1-WC-6DUP	-51	0.0368	-	7470A	PPB	20 May 2025 10:55:40	U	SMPL
Q2062-24MS - 1	L1-WC-6MS	20199	4.3571	-	7470A	PPB	20 May 2025 10:57:56	U	SMPL
Q2062-24MSD - 1	L1-WC-6MSD	19325	4.1707	-	7470A	PPB	20 May 2025 11:00:12	U	SMPL
PB168071BL - 1	PBW	-214	0.0020	-	7470A	PPB	20 May 2025 11:02:32	U	SMPL
PB168071BS - 1	LCSW	16686	3.6076	-	7470A	PPB	20 May 2025 11:04:50	U	SMPL
Q2033-01 - 1	TW-WTS-08	-135	0.0189	-	7470A	PPB	20 May 2025 11:07:06	U	SMPL
CCV43 - 1	CCV43	24825	5.3441	-	7470A	PPB	20 May 2025 11:09:23	U	SMPL
CCB43 - 1	CCB43	-368	-0.0308	-	7470A	PPB	20 May 2025 11:11:38	U	SMPL
Q2064-01 - 1	RW8-SP100-70-20250514	-165	0.0125	-	7470A	PPB	20 May 2025 11:13:56	U	SMPL
Q2064-01DUP - 1	RW8-SP100-70-20250514DUP	-83	0.0300	-	7470A	PPB	20 May 2025 11:16:11	U	SMPL
Q2064-01MS - 1	RW8-SP100-70-20250514MS	15001	3.2481	-	7470A	PPB	20 May 2025 11:18:27	U	SMPL
Q2064-01MSD - 1	RW8-SP100-70-20250514MSD	17187	3.7145	-	7470A	PPB	20 May 2025 11:20:43	U	SMPL
Q2064-04 - 1	RW8-SP303-70-20250514	-201	0.0048	-	7470A	PPB	20 May 2025 11:23:01	U	SMPL
Q2064-05 - 1	RW8-SP100-90-20250514	-25	0.0424	-	7470A	PPB	20 May 2025 11:25:19	U	SMPL
Q2064-08 - 1	RW8-SP303-90-20250514	-40	0.0392	-	7470A	PPB	20 May 2025 11:27:35	U	SMPL
Q2065-01 - 1	VACTRUCK-728068	10765	2.3444	-	7470A	PPB	20 May 2025 11:29:52	U	SMPL
Q2067-01 - 1	303-PPR-FRAC	-347	-0.0263	-	7470A	PPB	20 May 2025 11:32:10	U	SMPL
Q2068-03 - 1	LAW-25-0074	101551	21.7135	-	7470A	PPB	20 May 2025 11:34:28	U	SMPL
CCV44 - 1	CCV44	22406	4.8280	-	7470A	PPB	20 May 2025 11:36:44	U	SMPL
CCB44 - 1	CCB44	-323	-0.0212	-	7470A	PPB	20 May 2025 11:39:22	U	SMPL
PB168033TB - 1	PB168033TB	-36	0.0400	-	7470A	PPB	20 May 2025 11:41:39	U	SMPL
Q2062-24LX5 - 1		16	0.0511	-	7470A	PPB	20 May 2025 11:43:55	U	SMPL
Q2062-24A - 1		18726	4.0429	-	7470A	PPB	20 May 2025 11:46:11	U	SMPL
Q2064-01LX5 - 1		-187	0.0078	-	7470A	PPB	20 May 2025 11:48:26	U	SMPL
Q2064-01A - 1		18438	3.9814	-	7470A	PPB	20 May 2025 11:50:44	U	SMPL
Q2068-03DLX5 - 1		21685	4.6742	-	7470A	PPB	20 May 2025 11:53:01	U	SMPL
CCV45 - 1	CCV45	22452	4.8378	-	7470A	PPB	20 May 2025 11:55:19	U	SMPL
CCB45 - 1	CCB45	-268	-0.0095	-	7470A	PPB	20 May 2025 11:57:39	U	SMPL