

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB135835

Slope : 99.1

BalanceID: WC SC-7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3178
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3191
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3161
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.05].

True Value of CCV1 = 2.00 Control Limits[+/- 0.05].

True Value of CCV2 = 12.00 Control Limits[+/- 0.05].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	05/20/2025	08:25
2	CAL2	1	Water	NA	NA	20.3	7.00	05/20/2025	08:26
3	CAL3	1	Water	NA	NA	20.3	10.02	05/20/2025	08:29
4	ICV	1	Water	NA	NA	20.2	7.01	05/20/2025	08:33
5	CCV1	1	Water	NA	NA	20.2	2.01	05/20/2025	08:35
6	Q2076-01	1	Solid	20.02	20	22.5	4.71	05/20/2025	08:45
7	Q2080-06	1	Solid	20.03	20	24.3	6.16	05/20/2025	09:00
8	Q2080-12	1	Solid	20.02	20	23.6	6.64	05/20/2025	09:10
9	Q2080-12DUP	1	Solid	20.03	20	23.7	6.65	05/20/2025	09:11
10	CCV2	1	Water	NA	NA	20.3	12.02	05/20/2025	09:15

WORKLIST(Hardcopy Internal Chain)

U3135835

WorkList Name : corrosivity q2080 WorkList ID : 189615 Department : Wet-Chemistry Date : 05-20-2025 07:54:31

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q2076-01	OILY-DEBRIS	Solid	Corrosivity	Cool 4 deg C	PSEG03	L31	05/19/2025	9045D
Q2080-06	PL-HRH-COMP-01	Solid	Corrosivity	Cool 4 deg C	PSEG03	L31	05/19/2025	9045D
Q2080-12	PL-HRH-COMP-02	Solid	Corrosivity	Cool 4 deg C	PSEG03	L31	05/19/2025	9045D

Date/Time 05/20/25 08:10
 Raw Sample Received by: sl wpc
 Raw Sample Relinquished by: CP SN

Date/Time 05/20/25 10:30
 Raw Sample Received by: CP SN
 Raw Sample Relinquished by: sl wpc