

LB131876


7470A

INSTRUMENT ID: CY1

Linear

 μ Abs.:

48693



A= 0.0000e+000

B= 2.0327e-004

C= 1.2611e-002

Rho= 0.9998347

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/01
0.0	0.000	-0.001	-0.001	-67	0.000	-67	0				1
0.2	0.200	0.203	0.003	939	0.0 %	939					2
2.5	2.500	2.427	-0.073	11879	0.0 %	11879					3
5.0	5.000	5.060	0.060	24829	0.0 %	24829					1
7.5	7.500	7.601	0.101	37330	0.0 %	37330					1
10.0	10.000	9.910	-0.090	48693	0.0 %	48693					1

LB135876 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 22 May 2025 10:52:08

Sample ID	Extended ID	μ Abs	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50-2	-67	-	0.00007470A	PPB		22 May 2025 11:07:45	S	Std
0.2 - 1	52.5	939	-	0.20007470A	PPB		22 May 2025 11:10:03	S	Std
2.5 - 1	35	11879	-	2.50007470A	PPB		22 May 2025 11:12:20	S	Std
5.0 - 1	57.5	24829	-	5.00007470A	PPB		22 May 2025 11:17:12	S	Std
7.5 - 1	51.5	37330	-	7.50007470A	PPB		22 May 2025 11:19:28	S	Std
10.0 - 1	51.5	48693	-	10.00007470A	PPB		22 May 2025 11:21:45	S	Std
ICV13 - 1	ICV13	20634	4.2068	-7470A	PPB		22 May 2025 11:27:18	U	SMPL
ICB13 - 1	ICB13	-205	-0.0291	-7470A	PPB		22 May 2025 11:29:32	U	SMPL
CCV50 - 1	CCV50	24541	5.0010	-7470A	PPB		22 May 2025 11:31:49	U	SMPL
CCB50 - 1	CCB50	-235	-0.0352	-7470A	PPB		22 May 2025 11:34:04	U	SMPL
CRA - 1	CRA	851	0.1856	-7470A	PPB		22 May 2025 11:36:23	U	SMPL
HighStd - 1	HighStd	47372	9.6417	-7470A	PPB		22 May 2025 11:38:38	U	SMPL
ChkStd - 1	ChkStd	32110	6.5395	-7470A	PPB		22 May 2025 11:40:53	U	SMPL
PB168123BL - 1	PBW	59	0.0246	-7470A	PPB		22 May 2025 11:43:12	U	SMPL
PB168123BS - 1	LCSW	22444	4.5747	-7470A	PPB		22 May 2025 11:45:30	U	SMPL
Q2071-04 - 1	L1-WC-7	349	0.0836	-7470A	PPB		22 May 2025 11:47:47	U	SMPL
Q2071-08 - 1	L1-WC-8	478	0.1098	-7470A	PPB		22 May 2025 11:50:05	U	SMPL
Q2071-12 - 1	L1-WC-9	587	0.1319	-7470A	PPB		22 May 2025 11:52:22	U	SMPL
Q2071-16 - 1	L2-WC-5	-77	-0.0030	-7470A	PPB		22 May 2025 11:54:39	U	SMPL
Q2071-20 - 1	L2-WC-6	-50	0.0024	-7470A	PPB		22 May 2025 11:56:54	U	SMPL
CCV51 - 1	CCV51	23453	4.7798	-7470A	PPB		22 May 2025 11:59:09	U	SMPL
CCB51 - 1	CCB51	-401	-0.0689	-7470A	PPB		22 May 2025 12:01:24	U	SMPL
Q2080-06 - 1	PL-HRH-COMP-01	97	0.0323	-7470A	PPB		22 May 2025 12:03:42	U	SMPL
Q2080-12 - 1	PL-HRH-COMP-02	364	0.0866	-7470A	PPB		22 May 2025 12:05:57	U	SMPL
Q2080-12DUP - 1	PL-HRH-COMP-02DUP	486	0.1114	-7470A	PPB		22 May 2025 12:08:13	U	SMPL
Q2080-12MS - 1	PL-HRH-COMP-02MS	21348	4.3519	-7470A	PPB		22 May 2025 12:10:28	U	SMPL
Q2080-12MSD - 1	PL-HRH-COMP-02MSD	17827	3.6362	-7470A	PPB		22 May 2025 12:12:44	U	SMPL
PB168067TB - 1	PB168067TB	312	0.0760	-7470A	PPB		22 May 2025 12:15:02	U	SMPL
Q2080-12LZX5 - 1		20	0.0167	-7470A	PPB		22 May 2025 12:17:20	U	SMPL
Q2080-12A - 1		20333	4.1456	-7470A	PPB		22 May 2025 12:19:38	U	SMPL
CCV52 - 1	CCV52	23307	4.7501	-7470A	PPB		22 May 2025 12:21:55	U	SMPL
CCB52 - 1	CCB52	-262	-0.0406	-7470A	PPB		22 May 2025 12:24:13	U	SMPL