

Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB135877

Slope : 98.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3178
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3191
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3161
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.05].

True Value of CCV1 = 2.00 Control Limits[+/- 0.05].

True Value of CCV2 = 12.00 Control Limits[+/- 0.05].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	05/22/2025	10:25
2	CAL2	1	Water	NA	NA	20.2	7.00	05/22/2025	10:26
3	CAL3	1	Water	NA	NA	20.3	10.02	05/22/2025	10:30
4	ICV	1	Water	NA	NA	20.3	7.00	05/22/2025	10:33
5	CCV1	1	Water	NA	NA	20.2	2.01	05/22/2025	10:35
6	Q2102-03	1	Water	NA	NA	20.3	6.29	05/22/2025	10:40
7	Q2102-03DUP	1	Water	NA	NA	20.4	6.30	05/22/2025	10:41
8	CCV2	1	Water	NA	NA	20.3	12.02	05/22/2025	10:45

WORKLIST(Hardcopy Internal Chain)

17135877

WorkList Name : ph w q2102 WorkList ID : 189704 Department : Wet-Chemistry Date : 05-22-2025 09:52:45

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q2102-03	LAW-25-0078	Water	pH	Cool 4 deg C	PSEG03	L21	05/21/2025	9040C

Date/Time 05/22/25 10:10
Raw Sample Received by: JF acc
Raw Sample Relinquished by: JD CSM

Date/Time 05/22/25 13:00
Raw Sample Received by: JD CSM
Raw Sample Relinquished by: JF acc