

Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB135898

Slope : 99.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3178
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3191
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3161
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.05].

True Value of CCV1 = 2.00 Control Limits[+/- 0.05].

True Value of CCV2 = 12.00 Control Limits[+/- 0.05].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	05/23/2025	07:35
2	CAL2	1	Water	NA	NA	20.2	7.00	05/23/2025	07:31
3	CAL3	1	Water	NA	NA	20.3	10.02	05/23/2025	07:36
4	ICV	1	Water	NA	NA	20.3	6.99	05/23/2025	07:39
5	CCV1	1	Water	NA	NA	20.2	2.01	05/23/2025	07:40
6	Q2110-01	1	Water	NA	NA	20.4	5.68	05/23/2025	07:55
7	Q2110-01DUP	1	Water	NA	NA	20.5	5.69	05/23/2025	07:56
8	Q2110-02	1	Water	NA	NA	20.1	5.47	05/23/2025	08:00
9	CCV2	1	Water	NA	NA	20.2	12.02	05/23/2025	09:01

WORKLIST(Hardcopy Internal Chain)

VB135898

WorkList Name : PH W Q2110 WorkList ID : 189716 Department : Wet-Chemistry Date : 05-23-2025 07:10:59

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q2110-01	A 60457	Water	pH	Cool 4 deg C	PSEG03	L41	05/22/2025	9040C
Q2110-02	A 60458	Water	pH	Cool 4 deg C	PSEG03	L41	05/22/2025	9040C

Date/Time 05/23/25 07:25
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]

Date/Time 05/23/25 12:00
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]