

LB135915


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

41736



A= 0.0000e+000

B= 2.3565e-004

C= 5.8581e-002

Rho= 0.9994698

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%RSD
0.0	0.000	0.016	0.016	-181	0.000	-181	0				1
0.2	0.200	0.222	0.022	695	0.0 %	695					1
2.5	2.500	2.314	-0.186	9569	0.0 %	9569					2
5.0	5.000	5.188	0.188	21768	0.0 %	21768					4
7.5	7.500	7.566	0.066	31858	0.0 %	31858					1
10.0	10.000	9.894	-0.106	41736	0.0 %	41736					1

LB135915 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 27 May 2025 11:21:03

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	SO	-181	-	0.0000	7470A	PPB	27 May 2025 11:27:47	S	Std
0.2 - 1	51.2	695	-	0.2000	7470A	PPB	27 May 2025 11:30:05	S	Std
2.5 - 1	52.5	9569	-	2.5000	7470A	PPB	27 May 2025 11:32:22	S	Std
5.0 - 1	55	21768	-	5.0000	7470A	PPB	27 May 2025 11:37:11	S	Std
7.5 - 1	57.5	31858	-	7.5000	7470A	PPB	27 May 2025 11:39:29	S	Std
10.0 - 1	57.5	41736	-	10.0000	7470A	PPB	27 May 2025 11:43:55	S	Std
ICV17 - 1	ICV17	17584	4.2023	-	7470A	PPB	27 May 2025 11:47:01	U	SMPL
ICB17 - 1	ICB17	-184	0.0152	-	7470A	PPB	27 May 2025 11:49:16	U	SMPL
CCV65 - 1	CCV65	22227	5.2965	-	7470A	PPB	27 May 2025 11:51:32	U	SMPL
CCB65 - 1	CCB65	-224	0.0058	-	7470A	PPB	27 May 2025 11:53:48	U	SMPL
CRA - 1	CRA	752	0.2358	-	7470A	PPB	27 May 2025 11:56:05	U	SMPL
HighStd - 1	HighStd	41705	9.8865	-	7470A	PPB	27 May 2025 11:58:21	U	SMPL
ChkStd - 1	ChkStd	29705	7.0587	-	7470A	PPB	27 May 2025 12:00:36	U	SMPL
PB168157BL - 1	PBW	-297	-0.0114	-	7470A	PPB	27 May 2025 12:02:55	U	SMPL
PB168157BS - 1	LCSW	17802	4.2537	-	7470A	PPB	27 May 2025 12:05:13	U	SMPL
Q2109-04 - 1	TP-02-MHF	-134	0.0270	-	7470A	PPB	27 May 2025 12:07:30	U	SMPL
Q2109-04DUP - 1	TP-02-MHFDUP	-29	0.0517	-	7470A	PPB	27 May 2025 12:09:48	U	SMPL
Q2109-04MS - 1	TP-02-MHFMS	15890	3.8031	-	7470A	PPB	27 May 2025 12:12:05	U	SMPL
Q2109-04MSD - 1	TP-02-MHFMSD	17016	4.0685	-	7470A	PPB	27 May 2025 12:14:23	U	SMPL
PB168168BL - 1	PBW	-179	0.0164	-	7470A	PPB	27 May 2025 12:16:40	U	SMPL
CCV66 - 1	CCV66	21924	5.2251	-	7470A	PPB	27 May 2025 12:19:00	U	SMPL
CCB66 - 1	CCB66	-142	0.0251	-	7470A	PPB	27 May 2025 12:21:15	U	SMPL
PB168168BS - 1	LCSW	14896	3.5689	-	7470A	PPB	27 May 2025 12:23:32	U	SMPL
Q2088-01 - 1	252820	1153	0.3303	-	7470A	PPB	27 May 2025 12:25:48	U	SMPL
Q2093-14 - 1	WATER-COMP	-280	-0.0074	-	7470A	PPB	27 May 2025 12:28:05	U	SMPL
Q2102-03 - 1	LAW-25-0078	-2	0.0581	-	7470A	PPB	27 May 2025 12:30:23	U	SMPL
Q2102-03DUP - 1	LAW-25-0078DUP	39	0.0678	-	7470A	PPB	27 May 2025 12:32:40	U	SMPL
Q2102-03MS - 1	LAW-25-0078MS	11268	2.7139	-	7470A	PPB	27 May 2025 12:34:56	U	SMPL
Q2102-03MSD - 1	LAW-25-0078MSD	10963	2.6421	-	7470A	PPB	27 May 2025 12:39:55	U	SMPL
Q2106-01 - 1		-90	0.0374	-	7470A	PPB	27 May 2025 12:42:12	U	SMPL
PB168129TB - 1	PB168129TB	-79	0.0400	-	7470A	PPB	27 May 2025 12:44:31	U	SMPL
CCV67 - 1	CCV67	21604	5.1496	-	7470A	PPB	27 May 2025 12:46:47	U	SMPL
CCB67 - 1	CCB67	-202	0.0110	-	7470A	PPB	27 May 2025 12:49:03	U	SMPL
Q2109-04LX5 - 1		56	0.0718	-	7470A	PPB	27 May 2025 12:51:21	U	SMPL
Q2109-04A - 1		17318	4.1396	-	7470A	PPB	27 May 2025 12:53:36	U	SMPL
Q2102-03LX5 - 1		-160	0.0209	-	7470A	PPB	27 May 2025 12:55:52	U	SMPL
Q2102-03A - 1		17351	4.1474	-	7470A	PPB	27 May 2025 12:58:09	U	SMPL
CCV68 - 1	CCV68	21333	5.0858	-	7470A	PPB	27 May 2025 13:00:25	U	SMPL
CCB68 - 1	CCB68	-293	-0.0105	-	7470A	PPB	27 May 2025 13:02:45	U	SMPL