

LB135927


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

3390E



A= 0.0000e+000

B= 2.9330e-004

C= -1.1068e-002

Rho= 0.9997256

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.050	-0.050	-132	0.000	-132	0				1
0.2	0.200	0.241	0.041	861	0.0 %	861					21
2.5	2.500	2.407	-0.093	8244	0.0 %	8244					14
5.0	5.000	5.166	0.166	17651	0.0 %	17651					3
7.5	7.500	7.502	0.002	25614	0.0 %	25614					10
10.0	10.000	9.934	-0.066	33906	0.0 %	33906					1

LB135927 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 28 May 2025 12:27:45

Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	SC	-132	-	0.0000	7470A	PPB	28 May 2025 12:42:46	S	Std
0.2 - 1	SC	861	-	0.2000	7470A	PPB	28 May 2025 12:45:04	S	Std
2.5 - 1	SC	8244	-	2.5000	7470A	PPB	28 May 2025 12:47:22	S	Std
5.0 - 1	SC	17651	-	5.0000	7470A	PPB	28 May 2025 12:54:54	S	Std
7.5 - 1	SC	25614	-	7.5000	7470A	PPB	28 May 2025 13:02:40	S	Std
10.0 - 1	SC	33906	-	10.0000	7470A	PPB	28 May 2025 13:13:13	S	Std
ICV01 - 1	ICV01	12549	3.6696	-	7470A	PPB	28 May 2025 13:22:24	U	SMPL
ICB01 - 1	ICB01	76	0.0112	-	7470A	PPB	28 May 2025 13:24:39	U	SMPL
CCV01 - 1	CCV01	16969	4.9660	-	7470A	PPB	28 May 2025 13:26:55	U	SMPL
CCB01 - 1	CCB01	33	-0.0014	-	7470A	PPB	28 May 2025 13:29:10	U	SMPL
CRA - 1	CRA	789	0.2203	-	7470A	PPB	28 May 2025 13:31:27	U	SMPL
HIGH STD - 1	HIGH STD	33137	9.7082	-	7470A	PPB	28 May 2025 13:33:42	U	SMPL
CHK STD - 1	CHK STD	23076	6.7572	-	7470A	PPB	28 May 2025 13:35:57	U	SMPL
PB168181BL - 1	PBW	-214	-0.0738	-	7470A	PPB	28 May 2025 13:38:15	U	SMPL
Q2127-10 - 1	COMP-1	-346	-0.1126	-	7470A	PPB	28 May 2025 13:42:50	U	SMPL
Q2127-20 - 1	COMP-2	-179	-0.0636	-	7470A	PPB	28 May 2025 13:45:08	U	SMPL
Q2128-04 - 1	TP03-MHM	-133	-0.0501	-	7470A	PPB	28 May 2025 13:47:25	U	SMPL
Q2130-02 - 1	TP-333333	-55	-0.0272	-	7470A	PPB	28 May 2025 13:49:42	U	SMPL
Q2132-02 - 1	AUD-25-80-81	23	-0.0043	-	7470A	PPB	28 May 2025 13:51:57	U	SMPL
CCV02 - 1	CCV02	17079	4.9983	-	7470A	PPB	28 May 2025 13:54:13	U	SMPL
CCB02 - 1	CCB02	54	0.0048	-	7470A	PPB	28 May 2025 13:56:28	U	SMPL
Q2132-02DUP - 1	AUD-25-80-81DUP	-191	-0.0671	-	7470A	PPB	28 May 2025 13:58:44	U	SMPL
Q2132-02MS - 1	AUD-25-80-81MS	14061	4.1131	-	7470A	PPB	28 May 2025 14:01:00	U	SMPL
Q2132-02MSD - 1	AUD-25-80-81MSD	14826	4.3375	-	7470A	PPB	28 May 2025 14:03:15	U	SMPL
PB168181BS - 1	LCSW	12287	3.5928	-	7470A	PPB	28 May 2025 14:05:48	U	SMPL
PB168158TB - 1	PB168158TB	-227	-0.0776	-	7470A	PPB	28 May 2025 14:08:30	U	SMPL
Q2132-02LX5 - 1		29	-0.0026	-	7470A	PPB	28 May 2025 14:10:46	U	SMPL
Q2132-02A - 1		14470	4.2330	-	7470A	PPB	28 May 2025 14:13:02	U	SMPL
CCV03 - 1	CCV03	16863	4.9349	-	7470A	PPB	28 May 2025 14:15:18	U	SMPL
CCB03 - 1	CCB03	30	-0.0023	-	7470A	PPB	28 May 2025 14:17:37	U	SMPL