

6135947

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH CONSULTING GROUP INC
284 Sheffield Street, Mountainside, NJ 07092

5/29/2025 12:39

Reviewed by : RM Instrument ID : Konelab

Test: CNEPA-NEW

Sample Id	Result	Dil. 1 +	Response	Errors
-----------	--------	----------	----------	--------

ICV001 ICV001	90.031	0.0	0.083	
---------------	--------	-----	-------	--

ICB001 ICB001	2.235	0.0	0.001	
---------------	-------	-----	-------	--

CCV001 CCV001	240.434	0.0	0.225	
---------------	---------	-----	-------	--

CCB001 CCB001	2.068	0.0	0.001	
---------------	-------	-----	-------	--

PB168202BL PBS202	1.999	0.0	0.001	
-------------------	-------	-----	-------	--

02117 03-MDL-SOIL-QT2-2025-03	6.054	0.0	0.005	
-------------------------------	-------	-----	-------	--

PB168203BL PBW203	2.183	0.0	0.001	
-------------------	-------	-----	-------	--

02117.01 03-MDL-WATER-QT2-2025-03	5.866	0.0	0.005	
-----------------------------------	-------	-----	-------	--

CCV002 CCV002	238.314	0.0	0.223	
---------------	---------	-----	-------	--

05/29/2025 CCB002 CCB002	1.955	0.0	0.001	
--------------------------	-------	-----	-------	--

RM

N	10
Mean	59.114
SD	98.8018
CV%	167.14

Aquakem v. 7.2AQ1

Results from time period:

Thu May 29 10:46:17 2025

Thu May 29 11:56:10 2025

Sample Id	Sam	Test	short	na	Te	Result	Result	Result date and time	Stat
S0.0	A	CNEPA-NEW	P			1.9375	µg/l	5/29/2025 10:46:17	
S5.0	A	CNEPA-NEW	P			6.1491	µg/l	5/29/2025 10:46:18	
S10.0	A	CNEPA-NEW	P			10.8404	µg/l	5/29/2025 10:46:19	
S100.0	A	CNEPA-NEW	P			95.3129	µg/l	5/29/2025 10:46:20	
S250.0	A	CNEPA-NEW	P			249.702	µg/l	5/29/2025 10:46:21	
S500.0	A	CNEPA-NEW	P			501.0581	µg/l	5/29/2025 10:46:22	
ICV001 ICV001	S	CNEPA-NEW	P			90.0309	µg/l	5/29/2025 11:22:41	
ICB001 ICB001	S	CNEPA-NEW	P			2.2353	µg/l	5/29/2025 11:22:43	
CCV001 CCV001	S	CNEPA-NEW	P			240.4345	µg/l	5/29/2025 11:22:45	
CCB001 CCB001	S	CNEPA-NEW	P			2.0681	µg/l	5/29/2025 11:22:48	
PB168202BL PBS202	S	CNEPA-NEW	P			1.9988	µg/l	5/29/2025 11:22:49	
Q2117-03-MDL-SOIL-QT2-2025-03	S	CNEPA-NEW	P			6.054	µg/l	5/29/2025 11:30:17	
PB168203BL PBW203	S	CNEPA-NEW	P			2.1827	µg/l	5/29/2025 11:30:19	
Q2117-01-MDL-WATER-QT2-2025-03	S	CNEPA-NEW	P			5.866	µg/l	5/29/2025 11:30:23	
CCV002 CCV002	S	CNEPA-NEW	P			238.3139	µg/l	5/29/2025 11:30:26	
CCB002 CCB002	S	CNEPA-NEW	P			1.9554	µg/l	5/29/2025 11:33:40	

CHEMTECH CONSULTING GROUP INC
284 Sheffield Street, Mountainside, NJ 07092

Reviewed by : RM

Instrument ID : Konelab

5/29/2025 10:49

Test CNEPA-NEW

Accepted

5/29/2025 10:49

Factor Slope
Bias y-intercept

1066 0.000938


-0.001

05/30/2025
RM

Coeff. of det.

0.999854

Errors



	Calibrator	Response	Calc. con.	Conc.	Re Errors
1	0.0PPBCN 5.0	0.001	1.9375	0.0000	
2	5.0PPBCN 5.0	0.005	6.1491	5.0000	23.0
3	10PPBCN 5.0	0.009	10.8404	10.0000	8.4
4	100PPBCN 5.0	0.088	95.3129	100.0000	-4.7
5	250PPBCN 5.0	0.233	249.7020	250.0000	-0.1
6	500PPBCN 5.0	0.469	501.0581	500.0000	0.2

05/29/2025
RM