

Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB135950

Analyst By : jignesh

Supervisor Review By : Iwona

Slope : 98.4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3178
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3191
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3161
pH 12.00 Buffer	W3200

True Value of ICV = 7.00 Control Limits[+/- 0.05].

True Value of CCV1 = 2.00 Control Limits[+/- 0.05].

True Value of CCV2 = 12.00 Control Limits[+/- 0.05].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	05/29/2025	16:00
2	CAL2	1	Water	NA	NA	20.2	7.00	05/29/2025	16:01
3	CAL3	1	Water	NA	NA	20.2	10.02	05/29/2025	16:03
4	ICV	1	Water	NA	NA	20.3	6.99	05/29/2025	16:05
5	CCV1	1	Water	NA	NA	20.2	2.01	05/29/2025	16:10
6	Q2137-03	1	Water	NA	NA	21.9	7.24	05/29/2025	16:20
7	Q2137-03DUP	1	Water	NA	NA	22.1	7.25	05/29/2025	16:21
8	Q2155-03	1	Water	NA	NA	20.9	6.78	05/29/2025	16:25
9	Q2155-03DUP	1	Water	NA	NA	21.1	6.80	05/29/2025	16:29
10	Q2155-04	1	Water	NA	NA	20.8	7.50	05/29/2025	16:35
11	Q2155-04DUP	1	Water	NA	NA	20.9	7.53	05/29/2025	16:40
12	Q2155-05	1	Water	NA	NA	20.4	7.14	05/29/2025	16:48
13	Q2155-05DUP	1	Water	NA	NA	20.5	7.15	05/29/2025	16:50
14	CCV2	1	Water	NA	NA	20.3	12.02	05/29/2025	16:55

WORKLIST(Hardcopy Internal Chain)

135950

WorkList Name : ph q2156

WorkList ID : 189831

Department : Wet-Chemistry

Date : 05-29-2025 15:29:31

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q2137-03	A MOO-25-0149	Water	pH	Cool 4 deg C	PSEG03	L31	05/28/2025	9040C
Q2155-03	C 447-WATER	Water	pH	Cool 4 deg C	PSEG03	L31	05/29/2025	9040C
Q2155-04	C 444	Water	pH	Cool 4 deg C	PSEG03	L31	05/29/2025	9040C
Q2155-05	A 445	Water	pH	Cool 4 deg C	PSEG03	L31	05/29/2025	9040C

Date/Time 05/29/25 15:40

Raw Sample Received by: [Signature]

Raw Sample Relinquished by: [Signature]

Date/Time

05/29/25

Raw Sample Received by: [Signature]

Raw Sample Relinquished by: [Signature]