

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB135953

Slope : 99.1

BalanceID: WC SC-7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3178
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3191
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3161
pH 12.00 Buffer	W3200

True Value of ICV = 7.00 Control Limits[+/- 0.05].

True Value of CCV1 = 2.00 Control Limits[+/- 0.05].

True Value of CCV2 = 12.00 Control Limits[+/- 0.05].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	05/30/2025	10:15
2	CAL2	1	Water	NA	NA	20.2	7.01	05/30/2025	10:16
3	CAL3	1	Water	NA	NA	20.3	10.02	05/30/2025	10:20
4	ICV	1	Water	NA	NA	20.3	7.00	05/30/2025	10:22
5	CCV1	1	Water	NA	NA	20.3	2.01	05/30/2025	10:25
6	Q2157-01	1	Solid	20.02	20	21.3	5.61	05/30/2025	10:30
7	Q2157-01DUP	1	Solid	20.03	20	21.5	5.63	05/30/2025	10:34
8	CCV2	1	Water	NA	NA	20.3	12.02	05/30/2025	10:35

WORKLIST(Hardcopy Internal Chain)

135953

WorkList Name : ph s q2157

WorkList ID : 189843

Department : Wet-Chemistry

Date : 05-30-2025 09:59:06

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q2157-01	291	Solid	pH	Cool 4 deg C	PSEG03	L31	05/29/2025	9045D

Date/Time 05/30/25 10:10

Raw Sample Received by: [Signature]

Raw Sample Relinquished by: [Signature]

Date/Time 05/30/25 12:10

Raw Sample Received by: [Signature]

Raw Sample Relinquished by: [Signature]