

LB135958


7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

40880



A= 0.0000e+000

B= 2.3688e-004

C= -7.8049e-002

Rho= 0.9977897

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%SD
0.0	0.000	-0.074	-0.074	15	0.000	15					
0.2	0.200	0.188	-0.012	1124	0.0 %	1124					
2.5	2.500	2.381	-0.119	10382	0.0 %	10382					
5.0	5.000	5.245	0.245	22473	0.0 %	22473					
7.5	7.500	7.854	0.354	33485	0.0 %	33485					
10.0	10.000	9.606	-0.394	40880	0.0 %	40880					

9/010
 ~6
 -5
 -5
 -5
 -4

LB135958 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 30 May 2025 11:19:59

Sample ID	Extended ID	Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	15	-	0.00007471B	PPB		30 May 2025 11:45:46	S	Std
0.2 - 1	51.2	1124	-	0.20007471B	PPB		30 May 2025 11:48:03	S	Std
2.5 - 1	52.5	10382	-	2.50007471B	PPB		30 May 2025 11:50:20	S	Std
5.0 - 1	55	22473	-	5.00007471B	PPB		30 May 2025 11:58:07	S	Std
7.5 - 1	59.5	33485	-	7.50007471B	PPB		30 May 2025 12:03:07	S	Std
10.0 - 1	510	40880	-	10.00007471B	PPB		30 May 2025 12:10:49	S	Std
ICV22 - 1	ICV22	16529	3.8374	-7471B	PPB		30 May 2025 12:16:18	U	SMPL
ICB22 - 1	ICB22	195	-0.0319	-7471B	PPB		30 May 2025 12:18:34	U	SMPL
CCV79 - 1	CCV79	22127	5.1634	-7471B	PPB		30 May 2025 12:20:52	U	SMPL
CCB79 - 1	CCB79	3	-0.0773	-7471B	PPB		30 May 2025 12:23:07	U	SMPL
CRA - 1	CRA	1134	0.1906	-7471B	PPB		30 May 2025 12:25:25	U	SMPL
HighStd - 1	HighStd	41087	9.6547	-7471B	PPB		30 May 2025 12:27:41	U	SMPL
ChkStd - 1	ChkStd	30182	7.0715	-7471B	PPB		30 May 2025 12:29:56	U	SMPL
PB168222BL - 1	PBS	147	-0.0432	-7471B	PPB		30 May 2025 12:35:01	U	SMPL
PB168222BS - 1	LCSS	15054	3.4880	-7471B	PPB		30 May 2025 12:42:02	U	SMPL
Q2136-01 - 1	OR-646-COMP-52	2937	0.6177	-7471B	PPB		30 May 2025 12:44:19	U	SMPL
Q2137-01 - 1	MOO-25-0151	540	0.0499	-7471B	PPB		30 May 2025 12:46:38	U	SMPL
Q2139-01 - 1	BELL-25-0010	-61	-0.0925	-7471B	PPB		30 May 2025 12:48:57	U	SMPL
Q2141-01 - 1	VNJ-215	860	0.1257	-7471B	PPB		30 May 2025 12:51:14	U	SMPL
Q2146-01 - 1	TP04-MHN-WC	1413	0.2567	-7471B	PPB		30 May 2025 12:53:30	U	SMPL
CCV80 - 1	CCV80	22392	5.2262	-7471B	PPB		30 May 2025 12:55:45	U	SMPL
CCB80 - 1	CCB80	58	-0.0643	-7471B	PPB		30 May 2025 12:58:00	U	SMPL
Q2150-01 - 1	TP-44	1981	0.3912	-7471B	PPB		30 May 2025 13:05:41	U	SMPL
Q2150-01DUP - 1	TP-44DUP	1787	0.3453	-7471B	PPB		30 May 2025 13:07:56	U	SMPL
Q2150-01MS - 1	TP-44MS	17621	4.0960	-7471B	PPB		30 May 2025 13:12:56	U	SMPL
Q2150-01MSD - 1	TP-44MSD	18342	4.2668	-7471B	PPB		30 May 2025 13:15:11	U	SMPL
Q2150-02 - 1	TP-42	1207	0.2079	-7471B	PPB		30 May 2025 13:17:29	U	SMPL
Q2150-03 - 1	TP-39	2894	0.6075	-7471B	PPB		30 May 2025 13:19:49	U	SMPL
Q2150-04 - 1	TP-48	639	0.0733	-7471B	PPB		30 May 2025 13:22:05	U	SMPL
Q2150-05 - 1	TP-47	772	0.1048	-7471B	PPB		30 May 2025 13:24:23	U	SMPL
Q2150-06 - 1	TP-50	2824	0.5909	-7471B	PPB		30 May 2025 13:26:40	U	SMPL
Q2150-07 - 1	TP-51	1885	0.3685	-7471B	PPB		30 May 2025 13:28:56	U	SMPL
CCV81 - 1	CCV81	21842	5.0959	-7471B	PPB		30 May 2025 13:31:13	U	SMPL
CCB81 - 1	CCB81	266	-0.0150	-7471B	PPB		30 May 2025 13:37:57	U	SMPL
Q2150-08 - 1	TP-52	3085	0.6527	-7471B	PPB		30 May 2025 13:40:13	U	SMPL
Q2150-09 - 1	TP-54	731	0.0951	-7471B	PPB		30 May 2025 13:42:28	U	SMPL
Q2150-10 - 1	TP-53	1175	0.2003	-7471B	PPB		30 May 2025 13:44:44	U	SMPL
Q2151-01 - 1	WC-1	2260	0.4573	-7471B	PPB		30 May 2025 13:47:00	U	SMPL
Q2155-01 - 1	446	387	0.0136	-7471B	PPB		30 May 2025 13:49:15	U	SMPL
Q2155-02 - 1	447-SOLID	254	-0.0179	-7471B	PPB		30 May 2025 13:51:33	U	SMPL
Q2157-01 - 1	291	152	-0.0420	-7471B	PPB		30 May 2025 13:53:49	U	SMPL
Q2150-01LX5 - 1		432	0.0243	-7471B	PPB		30 May 2025 13:56:06	U	SMPL
Q2150-01A - 1		19501	4.5414	-7471B	PPB		30 May 2025 13:58:23	U	SMPL
CCV82 - 1	CCV82	21694	5.0609	-7471B	PPB		30 May 2025 14:00:38	U	SMPL
CCB82 - 1	CCB82	169	-0.0380	-7471B	PPB		30 May 2025 14:02:57	U	SMPL