

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Eman

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB135960

Ambient Barometric Pressure (mmHg): 750.00

Thermometer ID: Flash Point

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3193

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	9	28.00	82.4	83.0	05/30/2025	08:45
2	Q2137-03		1	14	100.00	>212.0	>212.0	05/30/2025	09:15
3	Q2137-03DUP		1	14	100.00	>212.0	>212.0	05/30/2025	09:45
4	Q2143-03		1	15	100.00	>212.0	>212.0	05/30/2025	10:15

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06