

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Iwona

Parameter: Flash Point

Supervisor Review By: jignesh

Run Number: LB135963

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3193

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	29.00	84.2	83.9	06/02/2025	08:45
2	Q2155-03		1	12	100.00	>212.0	>212.0	06/02/2025	09:15
3	Q2155-03DUP		1	12	100.00	>212.0	>212.0	06/02/2025	09:45
4	Q2155-04		1	12	100.00	>212.0	>212.0	06/02/2025	10:15
5	Q2155-05		1	13	100.00	>212.0	>212.0	06/02/2025	10:45
6	Q2157-02		1	12	100.00	>212.0	>212.0	06/02/2025	11:15
7	Q2165-03		1	12	100.00	>212.0	>212.0	06/02/2025	11:45
8	Q2165-06		1	13	100.00	>212.0	>212.0	06/02/2025	12:15

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB 13 5963

WorkList Name : FP-6-02

WorkList ID : 189867

Department : Wet-Chemistry

Date : 06-02-2025 08:21:40

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q2155-03	447-WATER	Water	Flash Point	Cool 4 deg C	PSEG03	L31	05/29/2025	1010B
Q2155-04	444	Water	Flash Point	Cool 4 deg C	PSEG03	L31	05/29/2025	1010B
Q2155-05	445	Water	Flash Point	Cool 4 deg C	PSEG03	L31	05/29/2025	1010B
Q2157-02	289	Water	Flash Point	Cool 4 deg C	PSEG03	L31	05/29/2025	1010B
Q2165-03	TRE-25-63-64	Water	Flash Point	Cool 4 deg C	PSEG03	L31	05/30/2025	1010B
Q2165-06	TRE-25-65-66	Water	Flash Point	Cool 4 deg C	PSEG03	L31	05/30/2025	1010B

Date/Time 06/02/25 08:40
Raw Sample Received by: 12 (SC)
Raw Sample Relinquished by: 28 (OC)

Date/Time 06/02/25 13:00
Raw Sample Received by: 28 (OC)
Raw Sample Relinquished by: 12 (SC)