

7470A

Linear

$$\mu \text{ Abs.} :$$

3066E

A= 0.0000e+000

B= 3.1320e-004

C= -4.4145e-002

Rho= 0.9970648

Accept=Accepted

Max. Conc:— 10.000 PPB

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	
0.0	0.000	-0.108	-0.108	-203	0.000	-203	0				0/21
0.2	0.200	0.169	-0.031	680	0.0 %	680					-
2.5	2.500	2.374	-0.126	7720	0.0 %	7720					-16
5.0	5.000	5.419	0.419	17442	0.0 %	17442					-5
7.5	7.500	7.786	0.286	25001	0.0 %	25001					8
10.0	10.000	9.560	-0.440	30666	0.0 %	30666					.

0/21
-
-16
-5
8
4
-4

LB135970 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 02 Jun 2025 10:58:26

Sample ID	Extended ID	μ Abs	Conc.	Std. Conc	Method	Units	Date	Type	Type
0.0 - 1	80	-203	-	0.0000	7470A	PPB	02 Jun 2025 11:07:15	S	Std
0.2 - 1	50.2	680	-	0.2000	7470A	PPB	02 Jun 2025 11:09:32	S	Std
2.5 - 1	52.5	7720	-	2.5000	7470A	PPB	02 Jun 2025 11:11:50	S	Std
5.0 - 1	55	17442	-	5.0000	7470A	PPB	02 Jun 2025 11:19:31	S	Std
7.5 - 1	59.5	25001	-	7.5000	7470A	PPB	02 Jun 2025 11:28:10	S	Std
10.0 - 1	51V	30666	-	10.0000	7470A	PPB	02 Jun 2025 11:35:39	S	Std
ICV23 - 1	ICV23	11928	3.6917	-7470A	PPB	02 Jun 2025 11:39:01	U	SMPL	
ICB23 - 1	ICB23	31	-0.0344	-7470A	PPB	02 Jun 2025 11:44:10	U	SMPL	
CCV83 - 1	CCV83	17470	5.4274	-7470A	PPB	02 Jun 2025 11:46:27	U	SMPL	
CCB83 - 1	CCB83	-103	-0.0764	-7470A	PPB	02 Jun 2025 11:48:42	U	SMPL	
CRA - 1	CRA	658	0.1619	-7470A	PPB	02 Jun 2025 11:50:59	U	SMPL	
HighStd - 1	HighStd	31249	9.7430	-7470A	PPB	02 Jun 2025 11:53:14	U	SMPL	
ChkStd - 1	ChkStd	21257	6.6135	-7470A	PPB	02 Jun 2025 11:55:29	U	SMPL	
PB168230BL - 1	PBW	109	-0.0100	-7470A	PPB	02 Jun 2025 12:01:27	U	SMPL	
PB168230BS - 1	LCSW	13886	4.3049	-7470A	PPB	02 Jun 2025 12:06:25	U	SMPL	
Q2136-05 - 1	OR-646-COMP-52	-185	-0.1021	-7470A	PPB	02 Jun 2025 12:08:40	U	SMPL	
Q2143-01 - 1	ELI-46-36-25-58-53	-3	-0.0451	-7470A	PPB	02 Jun 2025 12:10:59	U	SMPL	
Q2143-02 - 1	ELI-57-43-35-26	-131	-0.0852	-7470A	PPB	02 Jun 2025 12:13:15	U	SMPL	
Q2144-01 - 1	OILY-DEBRIS-COMP-A	-248	-0.1218	-7470A	PPB	02 Jun 2025 12:15:33	U	SMPL	
Q2144-02 - 1	OILY-DEBRIS-COMP-B	-53	-0.0607	-7470A	PPB	02 Jun 2025 12:17:48	U	SMPL	
CCV84 - 1	CCV84	17004	5.2815	-7470A	PPB	02 Jun 2025 12:20:04	U	SMPL	
CCB84 - 1	CCB84	-22	-0.0510	-7470A	PPB	02 Jun 2025 12:22:19	U	SMPL	
Q2146-04 - 1	TP04-MHN-WC	-24	-0.0517	-7470A	PPB	02 Jun 2025 12:24:37	U	SMPL	
Q2149-01 - 1	FILTER-CAKE	9	-0.0413	-7470A	PPB	02 Jun 2025 12:26:52	U	SMPL	
Q2151-04 - 1	WC-1	48	-0.0291	-7470A	PPB	02 Jun 2025 12:29:07	U	SMPL	
Q2151-04DUP - 1	WC-1DUP	55	-0.0269	-7470A	PPB	02 Jun 2025 12:31:22	U	SMPL	
Q2151-04MS - 1	WC-1MS	12693	3.9313	-7470A	PPB	02 Jun 2025 12:36:15	U	SMPL	
Q2151-04MSD - 1	WC-1MSD	12457	3.8574	-7470A	PPB	02 Jun 2025 12:38:31	U	SMPL	
PB168190TB - 1	PB168190TB	-7	-0.0463	-7470A	PPB	02 Jun 2025 12:45:17	U	SMPL	
Q2151-04LX5 - 1		15	-0.0394	-7470A	PPB	02 Jun 2025 12:47:34	U	SMPL	
Q2151-04A - 1		14663	4.5483	-7470A	PPB	02 Jun 2025 12:52:45	U	SMPL	
CCV85 - 1	CCV85	16819	5.2235	-7470A	PPB	02 Jun 2025 12:55:00	U	SMPL	
CCB85 - 1	CCB85	144	0.0010	-7470A	PPB	02 Jun 2025 12:59:47	U	SMPL	