

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB135972

Slope : 98.6

BalanceID: WC SC-7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3178
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3191
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3161
pH 12.00 Buffer	W3200

True Value of ICV = 7.00 Control Limits[+/- 0.05].

True Value of CCV1 = 2.00 Control Limits[+/- 0.05].

True Value of CCV2 = 12.00 Control Limits[+/- 0.05].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	06/02/2025	12:10
2	CAL2	1	Water	NA	NA	20.2	7.01	06/02/2025	12:11
3	CAL3	1	Water	NA	NA	20.3	10.02	06/02/2025	12:15
4	ICV	1	Water	NA	NA	20.3	7.01	06/02/2025	12:20
5	CCV1	1	Water	NA	NA	20.2	2.02	06/02/2025	13:45
6	Q2175-01	1	Solid	20.02	20	20.3	6.72	06/02/2025	14:00
7	Q2175-01DUP	1	Solid	20.03	20	20.4	6.73	06/02/2025	14:01
8	CCV2	1	Water	NA	NA	20.3	12.02	06/02/2025	14:05

WORKLIST(Hardcopy Internal Chain)

135972

WorkList Name : ph s q2175 WorkList ID : 189870 Department : Wet-Chemistry Date : 06-02-2025 11:42:17

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q2175-01	C 32525	Solid	pH	Cool 4 deg C	PSEG03	L41	05/30/2025	9045D

Date/Time 06/02/25 12:00
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]

Date/Time 06/02/25 12:30
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]