

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Iwona

Parameter: Flash Point

Supervisor Review By: jignesh

Run Number: LB135975

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3193

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.1	06/02/2025	13:15
2	Q2131-01		1	5	47.00	116.6	116.3	06/02/2025	13:45
3	Q2131-03		1	12	100.00	>212.0	>212.0	06/02/2025	14:15
4	Q2131-03DUP		1	11	100.00	>212.0	>212.0	06/02/2025	14:45
5	Q2131-05		1	5	72.00	161.6	161.3	06/02/2025	15:15
6	Q2131-07		1	5	42.00	107.6	107.3	06/02/2025	15:45

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB135975

WorkList Name : fp-05-28

WorkList ID : 189774

Department : Wet-Chemistry

Date : 05-28-2025 08:15:00

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q2131-01	CAM-40624	Solid	Flash Point	Cool 4 deg C	PSEG03	L23	05/27/2025	1010B
Q2131-03	CAM-40625	Solid	Flash Point	Cool 4 deg C	PSEG03	L23	05/27/2025	1010B
Q2131-05	CAM-40626	Solid	Flash Point	Cool 4 deg C	PSEG03	L23	05/27/2025	1010B
Q2131-07	CAM-40627	Solid	Flash Point	Cool 4 deg C	PSEG03	L23	05/27/2025	1010B

Date/Time 06/02/25 13:05
 Raw Sample Received by: 12(201)
 Raw Sample Relinquished by: 28(201)

Date/Time 06/02/25 16:25
 Raw Sample Received by: 28(201)
 Raw Sample Relinquished by: 12(201)