

LB135987


7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

51509



A= 0.0000e+000

B= 1.9260e-004

C= 4.5926e-003

Rho= 0.9998223

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	6/6/0
0.0	0.000	0.002	0.002	-15	0.000	-15					
0.2	0.200	0.217	0.017	1104	0.0 %	1104					
2.5	2.500	2.465	-0.035	12775	0.0 %	12775					
5.0	5.000	4.951	-0.049	25684	0.0 %	25684					
7.5	7.500	7.640	0.140	39643	0.0 %	39643					
10.0	10.000	9.925	-0.075	51509	0.0 %	51509					

LB135987 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 03 Jun 2025 12:27:12

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-15	-	0.0000	7471B	PPB	03 Jun 2025 12:29:18	S	Std
0.2 - 1	50.2	1104	-	0.2000	7471B	PPB	03 Jun 2025 12:31:35	S	Std
2.5 - 1	52.5	12775	-	2.5000	7471B	PPB	03 Jun 2025 12:33:52	S	Std
5.0 - 1	55	25684	-	5.0000	7471B	PPB	03 Jun 2025 12:36:09	S	Std
7.5 - 1	57.5	39643	-	7.5000	7471B	PPB	03 Jun 2025 12:38:27	S	Std
10.0 - 1	510	51509	-	10.0000	7471B	PPB	03 Jun 2025 12:40:46	S	Std
ICV25 - 1	ICV25	21029	4.0547	-7471B	PPB	03 Jun 2025 12:43:46	U	SMPL	
ICB25 - 1	ICB25	-284	-0.0501	-7471B	PPB	03 Jun 2025 12:46:02	U	SMPL	
CCV90 - 1	CCV90	24460	4.7155	-7471B	PPB	03 Jun 2025 12:48:20	U	SMPL	
CCB90 - 1	CCB90	-272	-0.0478	-7471B	PPB	03 Jun 2025 12:50:35	U	SMPL	
CRA - 1	CRA	1052	0.2072	-7471B	PPB	03 Jun 2025 12:52:53	U	SMPL	
HighStd - 1	HighStd	51728	9.9673	-7471B	PPB	03 Jun 2025 12:55:08	U	SMPL	
ChkStd - 1	ChkStd	34428	6.6353	-7471B	PPB	03 Jun 2025 12:57:23	U	SMPL	
PB168248BL - 1	PBS	-196	-0.0332	-7471B	PPB	03 Jun 2025 12:59:48	U	SMPL	
PB168248BS - 1	LCSS	21026	4.0541	-7471B	PPB	03 Jun 2025 13:04:46	U	SMPL	
Q2152-01 - 1	OK-02-05292025	1532	0.2997	-7471B	PPB	03 Jun 2025 13:07:03	U	SMPL	
Q2153-01 - 1	TR-04-0592025	1008	0.1987	-7471B	PPB	03 Jun 2025 13:09:22	U	SMPL	
Q2159-01 - 1	TP05-MHO-WC	1035	0.2039	-7471B	PPB	03 Jun 2025 13:11:40	U	SMPL	
Q2160-01 - 1	TP04-MHG-WC	381	0.0780	-7471B	PPB	03 Jun 2025 13:13:57	U	SMPL	
Q2160-01DUP - 1	TP04-MHG-WCDUP	541	0.1088	-7471B	PPB	03 Jun 2025 13:18:29	U	SMPL	
CCV91 - 1	CCV91	23569	4.5439	-7471B	PPB	03 Jun 2025 13:20:44	U	SMPL	
CCB91 - 1	CCB91	-256	-0.0447	-7471B	PPB	03 Jun 2025 13:22:59	U	SMPL	
Q2160-01MS - 1	TP04-MHG-WCMS	20879	4.0258	-7471B	PPB	03 Jun 2025 13:25:17	U	SMPL	
Q2160-01MSD - 1	TP04-MHG-WCMSD	20591	3.9704	-7471B	PPB	03 Jun 2025 13:27:33	U	SMPL	
Q2160-05 - 1	TP05-MHH-WC	214	0.0458	-7471B	PPB	03 Jun 2025 13:29:51	U	SMPL	
Q2168-01 - 1	SAN-A1-A3	5519	1.0675	-7471B	PPB	03 Jun 2025 13:32:10	U	SMPL	
Q2168-05 - 1	SAN-B1-B3	7432	1.4360	-7471B	PPB	03 Jun 2025 13:34:25	U	SMPL	
Q2168-09 - 1	SAN-C1-C2	152491	29.3739	-7471B	PPB	03 Jun 2025 13:36:44	U	SMPL	
Q2172-01 - 1	TP06-MHQ	97	0.0233	-7471B	PPB	03 Jun 2025 13:39:03	U	SMPL	
Q2173-01 - 1	OR-400-CF-402B-COMP-23	12830	2.4756	-7471B	PPB	03 Jun 2025 13:41:51	U	SMPL	
Q2173-07 - 1	OR-400-CF-402B-COMP-24	18135	3.4973	-7471B	PPB	03 Jun 2025 13:44:09	U	SMPL	
Q2173-13 - 1	OR-400-CF-402B-COMP-25	7778	1.5026	-7471B	PPB	03 Jun 2025 13:46:27	U	SMPL	
CCV92 - 1	CCV92	23473	4.5254	-7471B	PPB	03 Jun 2025 13:48:45	U	SMPL	
CCB92 - 1	CCB92	-228	-0.0393	-7471B	PPB	03 Jun 2025 13:51:04	U	SMPL	
PB168249BL - 1	PBS	308	0.0639	-7471B	PPB	03 Jun 2025 13:53:23	U	SMPL	
PB168249BS - 1	LCSS	20880	4.0260	-7471B	PPB	03 Jun 2025 13:55:39	U	SMPL	
Q2175-02 - 1	52725	589	0.1180	-7471B	PPB	03 Jun 2025 13:57:55	U	SMPL	
Q2176-01 - 1	TP-46	643	0.1284	-7471B	PPB	03 Jun 2025 14:00:13	U	SMPL	
Q2176-02 - 1	TP-56	1839	0.3588	-7471B	PPB	03 Jun 2025 14:02:29	U	SMPL	
Q2176-03 - 1	TP-25	1912	0.3728	-7471B	PPB	03 Jun 2025 14:04:45	U	SMPL	
Q2176-04 - 1	TP-26	4289	0.8306	-7471B	PPB	03 Jun 2025 14:07:02	U	SMPL	
Q2176-05 - 1	TP-28	1179	0.2317	-7471B	PPB	03 Jun 2025 14:09:20	U	SMPL	
Q2176-06 - 1	TP-27	421	0.0857	-7471B	PPB	03 Jun 2025 14:11:39	U	SMPL	
Q2176-07 - 1	TP-31	974	0.1922	-7471B	PPB	03 Jun 2025 14:13:55	U	SMPL	
CCV93 - 1	CCV93	25339	4.8848	-7471B	PPB	03 Jun 2025 14:16:11	U	SMPL	
CCB93 - 1	CCB93	-185	-0.0310	-7471B	PPB	03 Jun 2025 14:18:27	U	SMPL	
Q2176-08 - 1	TP-65	770	0.1529	-7471B	PPB	03 Jun 2025 14:25:41	U	SMPL	
Q2176-08DUP - 1	TP-65DUP	757	0.1504	-7471B	PPB	03 Jun 2025 14:27:56	U	SMPL	
Q2176-08MS - 1	TP-65MS	19089	3.6811	-7471B	PPB	03 Jun 2025 14:30:12	U	SMPL	
Q2176-08MSD - 1	TP-65MSD	20776	4.0060	-7471B	PPB	03 Jun 2025 14:32:28	U	SMPL	
Q2177-02 - 1	B-187-SB01	8915	1.7216	-7471B	PPB	03 Jun 2025 14:34:47	U	SMPL	
Q2177-04 - 1	B-187-SB02	3319	0.6438	-7471B	PPB	03 Jun 2025 14:37:06	U	SMPL	
Q2177-06 - 1	B-202-SB01	5391	1.0429	-7471B	PPB	03 Jun 2025 14:39:25	U	SMPL	
Q2185-01 - 1	TP02-MHB-WC	811	0.1608	-7471B	PPB	03 Jun 2025 14:41:43	U	SMPL	
Q2185-05 - 1	TP01-MHB-WC	692	0.1379	-7471B	PPB	03 Jun 2025 14:44:02	U	SMPL	
Q2168-09DLX5 - 1		31709	6.1117	-7471B	PPB	03 Jun 2025 14:46:18	U	SMPL	
CCV94 - 1	CCV94	24689	4.7596	-7471B	PPB	03 Jun 2025 14:48:34	U	SMPL	
CCB94 - 1	CCB94	-313	-0.0557	-7471B	PPB	03 Jun 2025 14:51:03	U	SMPL	
Q2160-01LX5 - 1		356	0.0732	-7471B	PPB	03 Jun 2025 14:53:21	U	SMPL	
Q2160-01A - 1		20823	4.0150	-7471B	PPB	03 Jun 2025 14:55:38	U	SMPL	
Q2176-08LX5 - 1		113	0.0264	-7471B	PPB	03 Jun 2025 14:57:53	U	SMPL	
Q2176-08A - 1		21164	4.0807	-7471B	PPB	03 Jun 2025 15:00:12	U	SMPL	
CCV95 - 1	CCV95	24991	4.8178	-7471B	PPB	03 Jun 2025 15:02:27	U	SMPL	
CCB95 - 1	CCB95	-348	-0.0624	-7471B	PPB	03 Jun 2025 15:04:47	U	SMPL	