

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB135990

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: Flash Point

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3193

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	10	28.00	82.4	82.1	06/03/2025	13:40
2	Q2169-01		1	13	100.00	>212.0	>212.0	06/03/2025	14:10
3	Q2169-01DUP		1	13	100.00	>212.0	>212.0	06/03/2025	14:40
4	Q2169-03		1	15	100.00	>212.0	>212.0	06/03/2025	15:10
5	Q2170-01		1	15	100.00	>212.0	>212.0	06/03/2025	15:40

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

6135990

WorkList Name : fp6-3

WorkList ID : 189899

Department : Wet-Chemistry

Date : 06-03-2025 13:20:24

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q2169-01	303-PPR-1	Water	Flash Point	Cool 4 deg C	PSEG01	L41	05/30/2025	1010B
Q2169-03	303-PPR-2	Water	Flash Point	Cool 4 deg C	PSEG01	L41	05/30/2025	1010B
Q2170-01	51925-AQ	Water	Flash Point	Cool 4 deg C	PSEG03	L41	05/30/2025	1010B

Date/Time 06/03/2025 13:30
Raw Sample Received by: RM Lwin
Raw Sample Relinquished by: RM Lwin

Date/Time 06/03/2025 16:15
Raw Sample Received by: SB Lwin
Raw Sample Relinquished by: RM Lwin