

Analytical Summary Report

Analysis Method: 365.3
 Parameter: Phosphorus, Total
 Run Number: LB136009

ANALYST: Iwona
 SUPERVISOR REVIEW BY: jignesh
 BALANCE ID: WC SC-7

Reagent/Standard	Lot/Log #
calibration std. phosphate 1 ppm	WP113339
calibration std. phosphate 0.5 ppm	WP113338
calibration std. phosphate 0.3 ppm	WP113337
calibration std. phosphate 0.1 ppm	WP113336
calibration std. phosphate 0.05 ppm	WP113335
calibration std. 0 ppm	WP113334
phosphate CCV std.	WP113340
5N sulfuric acid	WP112831
Combined reagent	WP113376
Phenolphthalein indicator	WP113378
Sodium hydroxide, 1N	WP111323
Phosphate ICV-LCS Std	WP113341
Phosphate LOD-MDL Std 0.025ppm	WP113344

Intercept: -0.0032 Slope: 0.6559 Regression: 0.999778

Seq	Lab ID	True Value (mg/L)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	Result (mg/L)	%D	AnalDate	AnalTime
1	CAL1	0.00	1	50	50	0.000	0.005		06/04/2025	13:10
2	CAL2	0.05	1	50	50	0.031	0.052	4	06/04/2025	13:10
3	CAL3	0.10	1	50	50	0.065	0.104	4	06/04/2025	13:11
4	CAL4	0.30	1	50	50	0.183	0.284	-5.3	06/04/2025	13:11
5	CAL5	0.50	1	50	50	0.326	0.502	0.4	06/04/2025	13:12
6	CAL6	1.00	1	50	50	0.655	1.004	0.4	06/04/2025	13:12

Analytical Summary Report

Analysis Method: 365.3

ANALYST: Iwona

Parameter: Phosphorus, Total

SUPERVISOR REVIEW BY: jignesh

Run Number: LB136009

BALANCE ID: WC SC-7

Seq	Lab ID	True Value (mg/l)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	Result (mg/L)	AnalDate	AnalTime
1	ICV	0.50	1	50	50	0.321	0.494	06/04/2025	13:13
2	ICB		1	50	50	0.000	0.005	06/04/2025	13:13
3	CCV1	0.50	1	50	50	0.329	0.506	06/04/2025	13:30
4	CCB1		1	50	50	0.001	0.006	06/04/2025	13:30
5	RL Check	0.05	1	50	50	0.033	0.055	06/04/2025	13:31
6	PB168280BL		1	1	50	0.001	0.006	06/04/2025	13:31
7	PB168280BS	0.50	1	1	50	0.320	0.493	06/04/2025	13:32
8	Q2126-03		1	1	50	0.011	0.022	06/04/2025	13:32
9	CCV2	0.50	1	50	50	0.326	0.502	06/04/2025	13:33
10	CCB2		1	50	50	0.001	0.006	06/04/2025	13:33