

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB136010

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: Flash Point

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3193

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	9	28.00	82.4	82.1	06/04/2025	10:45
2	Q2175-04		1	14	100.00	>212.0	>212.0	06/04/2025	11:15
3	Q2175-04DUP		1	14	100.00	>212.0	>212.0	06/04/2025	11:45
4	Q2175-06		1	10	44.00	111.2	110.9	06/04/2025	12:15
5	Q2175-09		1	15	100.00	>212.0	>212.0	06/04/2025	12:45

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06