



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 6/6/2025

OVENTEMP IN Celsius(°C): 107
Time IN: 17:25
In Date: 06/05/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 104
Time OUT: 08:11
Out Date: 06/06/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID- OVEN

QC:LB136015

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
Q2207-01	BU-703-COMP-01	1	1.15	10.74	11.89	10.68	88.7	
Q2207-02	BU-703-VOC-01	2	1.18	10.23	11.41	10.18	88.0	
Q2207-03	BU-703-01A	3	1.19	10.01	11.2	9.78	85.8	
Q2207-04	BU-703-01B	4	1.15	10.52	11.67	10.14	85.5	
Q2207-05	BU-703-02A	5	1.19	10.02	11.21	10.31	91.0	
Q2207-06	BU-703-02B	6	1.15	10.23	11.38	10.37	90.1	
Q2207-07	BU-703-03A	7	1.15	10.03	11.18	10.25	90.7	
Q2207-08	BU-703-03B	8	1.14	10.73	11.87	10.68	88.9	
Q2207-10	BU-703-COMP-02	9	1.14	10.60	11.74	9.59	79.7	
Q2207-11	BU-703-VOC-02	10	1.13	10.85	11.98	11.06	91.5	
Q2207-12	BU-703-04A	11	1.16	10.64	11.8	10.26	85.5	
Q2207-13	BU-703-04B	12	1.15	10.82	11.97	10.35	85.0	
Q2207-14	BU-703-05A	13	1.11	10.46	11.57	8.7	72.6	
Q2207-15	BU-703-05B	14	1.14	10.79	11.93	8.84	71.4	
Q2207-16	BU-703-06A	15	1.19	10.16	11.35	8.63	73.2	
Q2207-17	BU-703-06B	16	1.17	10.51	11.68	8.85	73.1	
Q2207-19	BU-703-COMP-03	17	1.17	10.22	11.39	9.59	82.4	
Q2207-20	BU-703-VOC-03	18	1.19	10.32	11.51	10.92	94.3	
Q2207-21	BU-703-07A	19	1.14	10.84	11.98	10.61	87.4	
Q2207-22	BU-703-07B	20	1.18	10.37	11.55	9.6	81.2	
Q2207-23	BU-703-08A	21	1.19	10.56	11.75	9.24	76.2	
Q2207-24	BU-703-08B	22	1.18	10.24	11.42	8.76	74.0	
Q2207-25	BU-703-09A	23	1.19	10.23	11.42	9.77	83.9	
Q2207-26	BU-703-09B	24	1.18	10.17	11.35	9.77	84.5	
Q2207-28	BU-703-COMP-04	25	1.15	10.58	11.73	10.55	88.8	
Q2207-29	BU-703-VOC-04	26	1.14	10.93	12.07	11.01	90.3	
Q2207-30	BU-703-10A	27	1.15	10.40	11.55	10.66	91.4	
Q2207-31	BU-703-10B	28	1.14	10.67	11.81	10.8	90.5	



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 6/6/2025

OVENTEMP IN Celsius(°C): 107
Time IN: 17:25
In Date: 06/05/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 104
Time OUT: 08:11
Out Date: 06/06/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID- OVEN

QC:LB136015

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
Q2207-32	BU-703-11A	29	1.15	10.24	11.39	9.05	77.1	
Q2207-33	BU-703-11B	30	1.15	10.01	11.16	10.04	88.8	
Q2207-34	BU-703-12A	31	1.13	10.63	11.76	10.44	87.6	
Q2207-35	BU-703-12B	32	1.14	10.28	11.42	10.24	88.5	
Q2207-37	BU-703-COMP-05	33	1.13	10.19	11.32	8.67	74.0	
Q2207-38	BU-703-VOC-05	34	1.14	9.92	11.06	9.68	86.1	
Q2207-39	BU-703-13A	35	1.12	10.31	11.43	8.67	73.2	
Q2207-40	BU-703-13B	36	1.11	10.46	11.57	9.11	76.5	
Q2207-41	BU-703-14A	37	1.12	10.58	11.7	10.15	85.3	
Q2207-42	BU-703-14B	38	1.13	10.71	11.84	10.34	86.0	
Q2207-43	BU-703-15A	39	1.14	10.41	11.55	10.26	87.6	
Q2207-44	BU-703-15B	40	1.13	10.76	11.89	9.98	82.2	
Q2208-01	BU-703-COMP-06	41	1.15	10.59	11.74	10.11	84.6	
Q2208-02	BU-703-VOC-06	42	1.13	10.17	11.3	8.93	76.7	
Q2208-03	BU-703-16A	43	1.14	10.73	11.87	10.57	87.9	
Q2208-04	BU-703-16B	44	1.15	10.19	11.34	9.97	86.6	
Q2208-05	BU-703-17A	45	1.17	10.65	11.82	10.36	86.3	
Q2208-06	BU-703-17B	46	1.15	10.56	11.71	10.34	87.0	
Q2208-07	BU-703-18A	47	1.15	10.05	11.2	9.86	86.7	
Q2208-08	BU-703-18B	48	1.12	10.29	11.41	9.71	83.5	
Q2208-10	BU-703-COMP-07	49	1.13	10.28	11.41	10.02	86.5	
Q2208-11	BU-703-VOC-07	50	1.14	9.91	11.05	9.51	84.5	
Q2208-12	BU-703-19A	51	1.12	10.40	11.52	10.46	89.8	
Q2208-13	BU-703-19B	52	1.12	10.65	11.77	10.53	88.4	
Q2208-14	BU-703-20A	53	1.14	10.75	11.89	10.93	91.1	
Q2208-15	BU-703-20B	54	1.13	10.60	11.73	9.37	77.7	
Q2208-16	BU-703-21A	55	1.12	10.04	11.16	9.25	81.0	
Q2208-17	BU-703-21B	56	1.12	10.45	11.57	9.14	76.7	

PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 6/6/2025

OVENTEMP IN Celsius(°C): 107
Time IN: 17:25
In Date: 06/05/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 104
Time OUT: 08:11
Out Date: 06/06/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID- OVEN

QC:LB136015

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Sample Wt (g)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
Q2208-19	BU-703-COMP-08	57	1.13	10.20	11.33	10.28	89.7	
Q2208-20	BU-703-VOC-08	58	1.15	10.36	11.51	9.88	84.3	
Q2208-21	BU-703-22A	59	1.13	10.86	11.99	10.51	86.4	
Q2208-22	BU-703-22B	60	1.13	10.50	11.63	10.47	89.0	
Q2208-23	BU-703-23A	61	1.13	10.27	11.4	9.77	84.1	
Q2208-24	BU-703-23B	62	1.14	9.92	11.06	9.39	83.2	
Q2208-25	BU-703-24A	63	1.13	10.03	11.16	9.39	82.4	
Q2208-26	BU-703-24B	64	1.13	10.53	11.66	9.92	83.5	
Q2208-28	BU-703-COMP-09	65	1.14	10.54	11.68	9.73	81.5	
Q2208-29	BU-703-VOC-09	66	1.14	10.84	11.98	11.06	91.5	
Q2208-30	BU-703-25A	67	1.13	10.33	11.46	8.61	72.4	
Q2208-31	BU-703-25B	68	1.14	9.89	11.03	8.35	72.9	
Q2208-32	BU-703-26A	69	1.12	10.45	11.57	8.88	74.3	
Q2208-33	BU-703-26B	70	1.12	10.32	11.44	8.73	73.7	
Q2208-34	BU-703-27A	71	1.12	10.63	11.75	9.7	80.7	
Q2208-35	BU-703-27B	72	1.12	10.86	11.98	10.33	84.8	
Q2214-01	RBR251425	73	1.13	10.04	11.17	7.48	63.2	
Q2214-02	RBR251425-E2	74	1.13	10.10	11.23	7.95	67.5	
Q2218-01	72-11934	75	1.14	10.25	11.39	10.55	91.8	
Q2218-02	72-11934-E2	76	1.14	10.61	11.75	11.09	93.8	
Q2219-01	72-11995	77	1.15	10.36	11.51	10.5	90.3	
Q2219-02	72-11995-E2	78	1.14	9.94	11.08	9.63	85.4	
Q2220-01	52925	82	1.00	1.00	2.00	2.00	100.0	wipe sample
Q2221-01	34798	83	1.00	1.00	2.00	2.00	100.0	oil sample
Q2223-01	HR-01-06042025	84	1.14	10.44	11.58	11.02	94.6	
Q2223-02	HR-01-06042025-E2	85	1.14	10.47	11.61	10.5	89.4	
Q2223-03	HR-04-06042025	86	1.13	10.45	11.58	11.15	95.9	
Q2223-04	HR-04-06042025-E2	87	1.13	10.43	11.56	10.93	94.0	



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 6/6/2025

OVENTEMP IN Celsius(°C): 107
Time IN: 17:25
In Date: 06/05/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 104
Time OUT: 08:11
Out Date: 06/06/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID- OVEN

QC:LB136015

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
Q2224-01	EO-01-06042025	88	1.14	10.45	11.59	11.00	94.4	
Q2224-02	EO-01-06042025-E2	89	1.18	10.36	11.54	10.48	89.8	
Q2225-01	SU-03-06042025	90	1.14	10.84	11.98	11.17	92.5	
Q2225-02	SU-03-06042025-E2	91	1.13	10.47	11.6	10.52	89.7	
Q2226-01	TP06-MHI-WC	79	1.15	10.74	11.89	10.9	90.8	
Q2226-02	TP06-MHI-VOC	80	1.14	10.79	11.93	10.94	90.8	
Q2226-03	TP06-MHI-EPH	81	1.14	10.85	11.99	10.8	89.0	
Q2227-01	TP07-MHH-WC	92	1.15	9.94	11.09	9.25	81.5	
Q2227-02	TP07-MHH-VOC	93	1.19	10.66	11.85	10.3	85.5	
Q2227-03	TP07-MHH-EPH	94	1.19	10.06	11.25	9.55	83.1	
Q2228-01	TP08-MHI-WC	95	1.19	10.52	11.71	9.91	82.9	
Q2228-02	TP08-MHI-VOC	96	1.19	10.50	11.69	10.45	88.2	
Q2228-03	TP08-MHI-EPH	97	1.13	10.85	11.98	10.26	84.1	
Q2235-02	WC-A2-08-C	98	1.18	10.11	11.29	9.37	81.0	
Q2236-02	WC-A4-05A-C	99	1.18	10.14	11.32	9.07	77.8	
Q2236-06	WC-A2-04-C	100	1.16	9.97	11.13	8.31	71.7	
Q2236-10	WC-A2-05-C	101	1.19	10.17	11.36	9.24	79.2	
Q2236-14	WC-A2-06-C	102	1.14	9.97	11.11	9.43	83.1	
Q2236-18	WC-A2-07-C	103	1.19	10.33	11.52	8.67	72.4	
Q2238-01	0528	104	1.18	10.04	11.22	9.89	86.8	
Q2239-01	399-16TH-ST-COMP	105	1.00	1.00	2.00	2.00	100.0	debris
Q2240-01	TP-3	106	1.15	10.83	11.98	10.54	86.7	
Q2240-02	TP-3-EPH	107	1.14	9.89	11.03	9.06	80.1	
Q2240-03	TP-3-VOC	108	1.16	10.71	11.87	10.5	87.2	
Q2240-05	TP-2	109	1.15	10.51	11.66	9.84	82.7	
Q2240-06	TP-2-EPH	110	1.13	10.31	11.44	9.82	84.3	
Q2240-07	TP-2-VOC	111	1.14	10.21	11.35	9.6	82.9	
Q2240-09	TP-1	112	1.18	10.21	11.39	9.14	78.0	



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 6/6/2025

OVENTEMP IN Celsius(°C): 107
Time IN: 17:25
In Date: 06/05/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 104
Time OUT: 08:11
Out Date: 06/06/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID- OVEN

QC:LB136015

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
Q2240-10	TP-1-EPH	113	1.13	10.22	11.35	8.77	74.8	
Q2240-11	TP-1-VOC	114	1.18	10.43	11.61	10.34	87.8	
Q2241-01	TP-N	115	1.13	10.86	11.99	10.72	88.3	
Q2241-02	TP-N-VOC	116	1.14	10.84	11.98	11.37	94.4	
Q2241-03	TP-N-EPH	117	1.13	10.85	11.98	10.71	88.3	
Q2241-05	TP-S	118	1.11	10.49	11.6	10.66	91.0	
Q2241-06	TP-S-VOC	119	1.13	10.15	11.28	10.32	90.5	
Q2241-07	TP-S-EPH	120	1.14	10.13	11.27	10.28	90.2	
Q2242-01	TP09-MHJ	121	1.13	10.81	11.94	10.15	83.4	
Q2242-02	TP09-MHJ-VOC	122	1.12	10.56	11.68	10.87	92.3	
Q2242-03	TP09-MHJ-EPH	123	1.13	10.09	11.22	9.63	84.2	
Q2244-01	TP03-MHC	124	1.14	10.79	11.93	11.00	91.4	
Q2244-02	TP03-MHC-VOC	125	1.18	10.19	11.37	10.6	92.4	
Q2244-03	TP03-MHC-EPH	126	1.14	10.85	11.99	11.68	97.1	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$